

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV010419\
 Data File : VV009156.D
 Acq On : 04 Jan 2019 11:44
 Operator : SY/MD
 Sample : VSTD02564
 Misc : 5.00 G/10ML/MSVOA V/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD02564

Quant Time: Jan 05 04:23:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM010419S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jan 05 04:20:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.402	0.369	8.2	100	0.00
3 T	Chloromethane	0.272	0.250	8.1	100	0.00
4 S	Vinyl Chloride-d3	0.171	0.140	18.1	100	0.00
5 T	Vinyl chloride	0.291	0.268	7.9	100	0.00
6 T	Bromomethane	0.205	0.198	3.4	100	0.00
7 S	Chloroethane-d5	0.166	0.161	3.0	100	0.00
8 T	Chloroethane	0.180	0.192	-6.7	100	0.00
9 T	Trichlorofluoromethane	0.534	0.513	3.9	100	0.00
10 S	1,1-Dichloroethene-d2	0.572	0.488	14.7	100	0.00
11 T	1,1,2-Trichloro-1,2,2-trifl	0.356	0.335	5.9	100	0.00
12 T	1,1-Dichloroethene	0.309	0.295	4.5	100	0.00
13 T	Acetone	0.170	0.129	24.1	100	0.00
14 T	Carbon disulfide	0.882	0.830	5.9	100	0.00
15 T	Methyl Acetate	0.211	0.184	12.8	100	0.00
16 T	Methylene chloride	0.373	0.340	8.8	100	0.00
17 T	Methyl tert-butyl Ether	0.756	0.716	5.3	100	0.00
18 T	trans-1,2-Dichloroethene	0.327	0.320	2.1	100	0.00
19 T	1,1-Dichloroethane	0.543	0.527	2.9	100	0.00
20 S	2-Butanone-d5	0.120	0.091	24.2	100	0.00
21	2-Butanone	0.149	0.121	18.8	100	0.00
22 T	cis-1,2-Dichloroethene	0.350	0.346	1.1	100	0.00
23 T	Bromochloromethane	0.164	0.162	1.2	100	0.00
24 S	Chloroform-d	0.526	0.467	11.2	100	0.00
25 T	Chloroform	0.578	0.576	0.3	100	0.00
26 S	1,2-Dichloroethane-d4	0.282	0.241	14.5	100	0.00
27 T	1,2-Dichloroethane	0.388	0.368	5.2	100	0.00
28 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
29 S	Benzene-d6	1.114	0.988	11.3	100	0.00
30 T	Cyclohexane	0.554	0.524	5.4	100	0.00
31 T	1,1,1-Trichloroethane	0.608	0.604	0.7	100	0.00
32 T	Carbon tetrachloride	0.556	0.554	0.4	100	0.00
33 S	1,2-Dichloropropane-d6	0.319	0.277	13.2	100	0.00
34 T	Benzene	1.344	1.317	2.0	100	0.00
35 T	Trichloroethene	0.387	0.377	2.6	100	0.00
36 T	Methylcyclohexane	0.647	0.614	5.1	100	0.00
37 S	Toluene-d8	1.051	0.926	11.9	100	0.00
38 S	trans-1,3-Dichloropropene-d	0.152	0.127	16.4	100	0.00
39 S	2-Hexanone-d5	0.075	0.061	18.7	100	0.00
40 T	1,2-Dichloropropane	0.324	0.332	-2.5	100	0.00
41 T	Bromodichloromethane	0.444	0.444	0.0	100	0.00
42 T	cis-1,3-Dichloropropene	0.513	0.516	-0.6	100	0.00
43 T	4-Methyl-2-pentanone	0.309	0.259	16.2	100	0.00
44 T	Toluene	1.449	1.446	0.2	100	0.00
45 T	trans-1,3-Dichloropropene	0.449	0.441	1.8	100	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV010419\
 Data File : VV009156.D
 Acq On : 04 Jan 2019 11:44
 Operator : SY/MD
 Sample : VSTD02564
 Misc : 5.00 G/10ML/MSVOA V/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD02564

Quant Time: Jan 05 04:23:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM010419S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jan 05 04:20:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	1,1,2-Trichloroethane	0.286	0.279	2.4	100	0.00
47 T	Tetrachloroethene	0.335	0.330	1.5	100	0.00
48 S	1,1,2,2-Tetrachloroethane-d	0.351	0.300	14.5	100	0.00
49 T	2-Hexanone	0.227	0.191	15.9	100	0.00
50 T	Dibromochloromethane	0.352	0.349	0.9	100	0.00
51 T	1,2-Dibromoethane	0.299	0.288	3.7	100	0.00
52 T	Chlorobenzene	0.994	0.966	2.8	100	0.00
53 T	Ethylbenzene	1.639	1.620	1.2	100	0.00
54 T	m,p-Xylene	0.638	0.633	0.8	100	0.00
55 T	o-xylene	0.590	0.589	0.2	100	0.00
56 T	Styrene	0.996	1.011	-1.5	100	0.00
57 T	Isopropylbenzene	1.638	1.652	-0.9	100	0.00
58 T	1,1,2,2-Tetrachloroethane	0.384	0.362	5.7	100	0.00
59	1,2,3-Trichloropropane	0.307	0.280	8.8	100	0.00
60 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
61 S	1,2-Dichlorobenzene-d4	0.845	0.700	17.2	100	0.00
62 T	Bromoform	0.442	0.428	3.2	100	0.00
63 T	1,3-Dichlorobenzene	1.575	1.523	3.3	100	0.00
64 T	1,4-Dichlorobenzene	1.642	1.563	4.8	100	0.00
65 T	1,2-Dichlorobenzene	1.500	1.458	2.8	100	0.00
66 T	1,2-Dibromo-3-chloropropane	0.156	0.144	7.7	100	0.00
67	1,3,5-Trichlorobenzene	1.270	1.225	3.5	100	0.00
68 T	1,2,4-trichlorobenzene	1.034	0.992	4.1	100	0.00
69	Naphthalene	2.106	1.852	12.1	100	0.00
70 T	1,2,3-Trichlorobenzene	0.949	0.898	5.4	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0