



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV010422\
Data File : VV024262.D
Acq On : 04 Jan 2022 09:58
Operator : SY/MD
Sample : BFB259
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB259

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR010422WMA.M
Title : TRACE VOA SFAM1.0
Last Update : Tue Jan 04 18:34:35 2022



AutoFind: Scans 2809, 2810, 2811; Background Corrected with Scan 2796

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	17997	PASS
75	95	30	80	54.2	54653	PASS
95	95	100	100	100.0	100920	PASS
96	95	5	9	6.7	6731	PASS
173	174	0.00	2	0.5	466	PASS
174	95	50	120	99.9	100864	PASS
175	174	5	9	7.5	7526	PASS
176	174	95	105	97.2	98021	PASS
177	176	5	10	6.6	6444	PASS

m/z	Abundance
43.90	111.0

Instrument :
MSVOA_V
ClientSampleId :
BFB259

m/z	Abundance
36.00	1084.0
37.10	4845.0
38.10	4419.0
39.10	1695.0
39.80	133.0
44.10	729.0
45.00	912.0
46.10	151.0
47.00	786.0
48.00	638.0
49.10	3755.0
50.10	18320.0
51.10	5707.0
52.10	326.0
55.00	296.0
56.10	1474.0
57.10	2974.0
58.00	103.0
60.00	1226.0
61.10	5439.0
62.10	5301.0
63.10	4163.0
64.10	624.0
65.00	591.0
65.90	140.0
67.00	417.0
68.00	10941.0
69.10	10775.0
70.10	787.0
72.00	679.0
73.00	5331.0
74.10	19424.0
75.10	55272.0
76.10	4795.0
77.00	515.0
77.90	339.0
79.00	4489.0
80.00	1254.0
81.00	4438.0
81.90	1059.0
82.90	128.0
85.90	112.0
87.00	3710.0
88.00	4006.0
91.00	577.0
92.00	3773.0
93.00	5033.0
94.10	12592.0
95.10	102640.0
96.10	7102.0
97.10	156.0
103.20	133.0
103.90	613.0
104.90	170.0
105.90	669.0
106.90	131.0
111.00	148.0
112.90	186.0
115.00	197.0
115.90	582.0
116.90	910.0
118.00	577.0
119.00	738.0
124.10	138.0
127.90	490.0
128.90	216.0
129.90	549.0
130.90	263.0
134.90	246.0
137.00	347.0
139.90	107.0
140.90	1576.0
141.80	299.0
142.90	1743.0
145.00	342.0
145.90	326.0
147.00	119.0
147.90	425.0
148.90	132.0
150.00	163.0
152.80	101.0
153.80	143.0
155.00	255.0
157.00	156.0
158.80	256.0
160.90	224.0
171.10	238.0
172.00	939.0
174.00	100624.0
175.00	7779.0
176.00	96488.0
177.00	6361.0

Instrument :
MSVOA_V
ClientSampleId :
BFB259

178.00

213.0

Instrument :
MSVOA_V
ClientSampleId :
BFB259

m/z	Abundance
36.10	1089.0
37.10	4730.0
38.10	4392.0
39.10	1639.0
39.90	102.0
44.10	690.0
45.00	920.0
46.10	131.0
47.00	771.0
48.00	621.0
49.10	4032.0
50.10	18560.0
51.10	5593.0
52.10	290.0
55.00	321.0
56.10	1497.0
57.10	3056.0
60.00	1309.0
61.10	5638.0
62.00	5460.0
63.10	4109.0
64.00	585.0
65.00	540.0
67.00	421.0
68.00	11009.0
69.10	10958.0
70.10	731.0
72.00	729.0
73.00	5364.0
74.10	19776.0
75.10	56576.0
76.10	4710.0
76.90	568.0
77.80	341.0
79.00	4540.0
80.00	1335.0
81.00	4609.0
81.90	1043.0
82.90	158.0
87.00	3564.0
88.00	4056.0
91.00	532.0
92.00	3938.0
93.00	5203.0
94.10	13008.0
95.10	104432.0
96.10	6917.0
97.10	157.0
103.00	104.0
104.00	676.0
104.80	168.0
105.90	727.0
106.90	147.0
111.00	146.0
112.00	111.0
112.90	156.0
115.00	228.0
115.90	616.0
116.90	956.0
117.90	589.0
119.00	771.0
124.10	139.0
127.90	497.0
128.90	199.0
129.90	486.0
130.90	246.0
134.90	243.0
137.00	346.0
139.80	134.0
140.90	1599.0
141.80	239.0
142.90	1822.0
143.80	100.0
145.00	361.0
145.90	298.0
146.90	101.0
147.90	406.0
148.90	112.0
149.90	201.0
152.80	123.0
153.90	177.0
154.90	324.0
156.90	192.0
158.80	259.0
160.90	207.0
171.00	217.0
172.00	919.0
174.00	104264.0
175.00	7796.0
176.00	101288.0
177.00	6623.0
178.00	204.0

Instrument :
MSVOA_V
ClientSampleId :
BFB259

m/z	Abundance
36.10	857.0
37.10	4117.0
38.10	3872.0
39.10	1510.0
40.00	127.0
44.10	654.0
45.00	882.0
46.10	124.0
47.00	694.0
48.00	553.0
49.10	3905.0
50.10	17112.0
51.10	5018.0
52.00	246.0
55.00	296.0
56.00	1397.0
57.00	2738.0
60.00	1154.0
61.10	5108.0
62.00	5022.0
63.10	3635.0
64.00	487.0
65.10	524.0
67.00	328.0
68.00	10137.0
69.10	10158.0
70.00	695.0
72.00	697.0
73.00	4852.0
74.10	18048.0
75.10	52112.0
76.10	4183.0
76.90	529.0
77.80	300.0
79.00	3945.0
80.00	1186.0
81.00	4258.0
81.90	960.0
82.90	183.0
87.00	3121.0
88.00	3670.0
90.90	442.0
92.00	3560.0
93.00	4837.0
94.10	12492.0
95.10	95688.0
96.10	6174.0
97.00	159.0
104.00	635.0
104.80	141.0
105.90	652.0
106.90	158.0
111.10	114.0
112.00	113.0
112.80	114.0
115.00	205.0
115.90	517.0
116.90	941.0
118.00	541.0
119.00	740.0
124.10	130.0
127.90	459.0
128.80	181.0
129.90	415.0
131.00	221.0
134.90	227.0
136.90	293.0
139.80	148.0
140.90	1588.0
141.80	164.0
143.00	1694.0
143.80	109.0
145.00	359.0
145.90	213.0
148.00	348.0
149.90	192.0
152.80	130.0
153.90	157.0
154.90	358.0
156.90	214.0
158.80	217.0
161.20	172.0
171.00	158.0
172.00	846.0
173.00	1399.0
174.00	97704.0
175.00	7002.0
176.00	96288.0
177.00	6349.0
178.00	190.0

Instrument :
MSVOA_V
ClientSampleId :
BFB259