

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV010521\
 Data File : VV019886.D
 Acq On : 05 Jan 2021 10:46
 Operator : SY/MD
 Sample : VV0105WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK67

Quant Time: Jan 06 07:16:45 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 06 07:08:24 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	556883	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	528656	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.26	152	243617	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	194680	45.18	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.36%
7) Chloroethane-d5	1.57	69	165524	48.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.64%
11) 1,1-Dichloroethene-d2	2.11	63	296581	36.41	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.82%
21) 2-Butanone-d5	3.89	46	311660	97.57	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.57%
24) Chloroform-d	4.35	84	379410	46.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.20%
26) 1,2-Dichloroethane-d4	5.04	65	268068	50.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.74%
32) Benzene-d6	5.05	84	754930	49.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.18%
36) 1,2-Dichloropropane-d6	6.07	67	244868	48.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.84%
41) Toluene-d8	7.32	98	653677	48.24	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.48%
43) trans-1,3-Dichloropropene-	7.62	79	113884	45.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.54%
47) 2-Hexanone-d5	8.09	63	175393	84.78	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.78%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	291634	44.80	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.60%
64) 1,2-Dichlorobenzene-d4	11.63	152	238661	51.22	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.44%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV010521\
Data File : VV019886.D
Acq On : 05 Jan 2021 10:46
Operator : SY/MD
Sample : VV0105WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK67

Quant Time: Jan 06 07:16:45 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 06 07:08:24 2021
Response via : Initial Calibration

