


Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VW010521\
Data File : VW019888.D
Acq On : 05 Jan 2021 11:46
Operator : SY/MD
Sample : M1016-02
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG4Q7

Manual Integrations
APPROVED

MMDadoda
1/6/2021 3:31:28 PM

Quant Time: Jan 06 07:27:46 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 06 07:08:24 2021
Response via : Initial Calibration



Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV010521\
 Data File : VV019888.D
 Acq On : 05 Jan 2021 11:46
 Operator : SY/MD
 Sample : M1016-02
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampled :
 BG4Q7

Manual Integrations
 APPROVED

MMDadoda
 1/6/2021 3:31:28 PM

Quant Time: Jan 06 07:11:02 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 06 07:08:24 2021
 Response via : Initial Calibration



TIC: VV019888.D

(24) Chloroform-d (S)

4.482min (+0.132) 0.02ug/L

response 163

Ion	Exp%	Act%
84.00	100	100
86.00	61.60	51.53
47.00	46.50	58.28
49.00	28.00	28.22

Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV010521\
 Data File : VV019888.D
 Acq On : 05 Jan 2021 11:46
 Operator : SY/MD
 Sample : M1016-02
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampled :
 BG4Q7

Manual Integrations
 APPROVED

MMDadoda
 1/6/2021 3:31:28 PM

Quant Time: Jan 06 07:11:02 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 06 07:08:24 2021
 Response via : Initial Calibration



TIC: VV019888.D

(24) Chloroform-d (S)

4.354min (+0.003) 48.12ug/L m

response 393024

Ion	Exp%	Act%
84.00	100	100
86.00	61.60	0.02#
47.00	46.50	0.02#
49.00	28.00	0.01#

01/18/21 01/18/21 Cy

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV010521\
 Data File : VV019888.D
 Acq On : 05 Jan 2021 11:46
 Operator : SY/MD
 Sample : M1016-02
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG4Q7

Manual Integrations
 APPROVED

MMDadoda
 1/6/2021 3:31:28 PM

Quant Time: Jan 06 07:27:46 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 06 07:08:24 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	558640	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	532799	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.26	152	241148	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	195327	45.18	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.36%
7) Chloroethane-d5	1.57	69	173158	50.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.78%
11) 1,1-Dichloroethene-d2	2.11	63	308481	37.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.50%
21) 2-Butanone-d5	3.89	46	312134	97.42	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.42%
24) Chloroform-d	4.35	84	393024m	48.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.24%
26) 1,2-Dichloroethane-d4	5.04	65	274248	51.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.74%
32) Benzene-d6	5.05	84	769620	50.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.32%
36) 1,2-Dichloropropane-d6	6.07	67	250177	49.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.18%
41) Toluene-d8	7.32	98	674481	49.39	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.78%
43) trans-1,3-Dichloropropene-	7.63	79	114727	45.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.50%
47) 2-Hexanone-d5	8.09	63	164994	79.13	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.13%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	284154	43.31	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.62%
64) 1,2-Dichlorobenzene-d4	11.63	152	245511	53.23	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.46%

0.118/214

Target Compounds				Ovalue
13) Acetone	2.18	43	15215	6.134 ug/L 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed