

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW010522\
 Data File : VW024277.D
 Acq On : 05 Jan 2022 14:37
 Operator : SY/MD
 Sample : VW0105WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK288

Quant Time: Jan 06 01:00:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 06 00:59:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	278937	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	263158	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	130731	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	91240	45.450	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.900%
7) Chloroethane-d5	1.571	69	71533	46.907	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.820%
11) 1,1-Dichloroethene-d2	2.111	63	136717	36.439	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.880%
21) 2-Butanone-d5	3.889	46	100596	89.942	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.940%
24) Chloroform-d	4.349	84	167820	45.740	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.480%
26) 1,2-Dichloroethane-d4	5.031	65	113232	46.096	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.200%
32) Benzene-d6	5.050	84	324905	47.312	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.620%
36) 1,2-Dichloropropane-d6	6.069	67	90640	47.096	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.200%
41) Toluene-d8	7.317	98	312005	46.925	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.840%
43) trans-1,3-Dichloroprop...	7.622	79	50271	45.361	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.720%
47) 2-Hexanone-d5	8.088	63	68264	81.640	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.640%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	113702	41.340	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.680%
66) 1,2-Dichlorobenzene-d4	11.625	152	131085	50.305	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.620%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV010522\
Data File : VV024277.D
Acq On : 05 Jan 2022 14:37
Operator : SY/MD
Sample : VV0105WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK288

Quant Time: Jan 06 01:00:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 06 00:59:13 2022
Response via : Initial Calibration

