

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV010621\
 Data File : VV019906.D
 Acq On : 06 Jan 2021 12:08
 Operator : SY/MD
 Sample : VSTD00232
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Jan 07 06:26:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM010621.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jan 07 06:25:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	8560	0.50	ug/L	# 0.00
5) Chlorobenzene-d5	8.863	117	8346	0.50	ug/L	# 0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	4118	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	13204	1.56	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	13549	1.23	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	16616	1.70	ug/L	0.00
8) Toluene-d8	7.317	98	37159	1.94	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	9694	1.05	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	25077	2.12	ug/L	97
6) Trichloroethene	5.925	95	16868	1.68	ug/L	97
9) 1,2-Dibromoethane	8.353	107	9490	1.37	ug/L #	95
12) 1,2,3-Trichloropropane	10.267	75	9233	1.49	ug/L #	70
13) 1,2-Dibromo-3-chloropr...	12.426	75	1231	0.97	ug/L	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV010621\
Data File : VV019906.D
Acq On : 06 Jan 2021 12:08
Operator : SY/MD
Sample : VSTD00232
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Jan 07 06:26:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM010621.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jan 07 06:25:13 2021
Response via : Initial Calibration









#3
 Vinyl chloride
 Concen: 2.12 ug/L
 RT: 1.310 min Scan# 59
 Delta R.T. 0.000 min
 Lab File: VV019906.D
 Acq: 06 Jan 2021 12:08




Instrument :
 MSVOA_V
 ClientSampled :


Tgt Ion: 62 Resp: 25077
 Ion Ratio Lower Upper
 62 100
 64 32.7 24.3 45.1



#6
 Trichloroethene
 Concen: 1.68 ug/L
 RT: 5.925 min Scan# 418
 Delta R.T. 0.000 min
 Lab File: VV019906.D
 Acq: 06 Jan 2021 12:08


Tgt Ion: 95 Resp: 16868
 Ion Ratio Lower Upper
 95 100
 97 65.7 47.3 87.9
 132 97.7 65.8 122.2
 130 100.5 68.6 127.4








#9
1,2-Dibromoethane
Concen: 1.37 ug/L
RT: 8.353 min Scan# 545
Delta R.T. -0.015 min
Lab File: VV019906.D
Acq: 06 Jan 2021 12:08


Instrument :
MSVOA_V
ClientSampled :





Tgt Ion: 107 Resp: 9490
Ion Ratio Lower Upper
107 100
109 91.4 66.6 123.6
188 4.6 8.6 12.8#




#12
1,2,3-Trichloropropane
Concen: 1.49 ug/L
RT: 10.267 min Scan# 659
Delta R.T. 0.000 min
Lab File: VV019906.D
Acq: 06 Jan 2021 12:08




Tgt Ion: 75 Resp: 9233
Ion Ratio Lower Upper
75 100
110 38.2 0.2 0.2#
77 32.2 0.0 0.0#





#13
 1,2-Dibromo-3-chloropropane
 Concen: 0.97 ug/L
 RT: 12.426 min Scan# 785
 Delta R.T. 0.000 min
 Lab File: VV019906.D
 Acq: 06 Jan 2021 12:08

Instrument :
 MSVOA_V
 ClientSampleId :



Tgt Ion: 75 Resp: 1231

Ion	Ratio	Lower	Upper
75	100		
157	75.3	60.6	90.8
155	63.8	55.8	83.6

