

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV010621\
 Data File : VV019904.D
 Acq On : 06 Jan 2021 11:22
 Operator : SY/MD
 Sample : VSTD0.536
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5202

Quant Time: Jan 07 06:45:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM010621.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jan 07 06:31:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	8282	0.50	ug/L	# 0.00
5) Chlorobenzene-d5	8.863	117	7933	0.50	ug/L	# 0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	4045	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	3375	0.49	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	3409	0.50	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	4160	0.51	ug/L	0.00
8) Toluene-d8	7.317	98	9119	0.49	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	2429	0.51	ug/L	0.00
Target Compounds						Qvalue
3) Vinyl chloride	1.310	62	6222	0.50	ug/L	100
6) Trichloroethene	5.925	95	4139	0.51	ug/L	100
9) 1,2-Dibromoethane	8.368	107	2308	0.53	ug/L	100
12) 1,2,3-Trichloropropane	10.267	75	2242	0.51	ug/L #	70
13) 1,2-Dibromo-3-chloropr...	12.426	75	294	0.48	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV010621\
Data File : VV019904.D
Acq On : 06 Jan 2021 11:22
Operator : SY/MD
Sample : VSTD0.536
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5202

Quant Time: Jan 07 06:45:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM010621.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jan 07 06:31:40 2021
Response via : Initial Calibration

