

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV010621\
 Data File : VV019906.D
 Acq On : 06 Jan 2021 12:08
 Operator : SY/MD
 Sample : VSTD00232
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD002204

Quant Time: Jan 07 06:46:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM010621.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jan 07 06:31:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	8560	0.50	ug/L	# 0.00
5) Chlorobenzene-d5	8.863	117	8346	0.50	ug/L	# 0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	4118	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	13204	1.84	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	13549	1.92	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	16616	1.95	ug/L	0.00
8) Toluene-d8	7.317	98	37159	1.91	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	9694	1.92	ug/L	0.00
Target Compounds						Qvalue
3) Vinyl chloride	1.310	62	25077	1.97	ug/L	97
6) Trichloroethene	5.925	95	16868	1.96	ug/L	97
9) 1,2-Dibromoethane	8.353	107	9490	2.06	ug/L #	95
12) 1,2,3-Trichloropropane	10.267	75	9233	2.05	ug/L #	70
13) 1,2-Dibromo-3-chloropr...	12.426	75	1231	1.96	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV010621\
Data File : VV019906.D
Acq On : 06 Jan 2021 12:08
Operator : SY/MD
Sample : VSTD00232
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002204

Quant Time: Jan 07 06:46:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM010621.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jan 07 06:31:40 2021
Response via : Initial Calibration

