

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV010621\
 Data File : VV019907.D
 Acq On : 06 Jan 2021 12:35
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV205

Quant Time: Jan 07 06:48:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM010621.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jan 07 06:31:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	8169	0.50	ug/L	# 0.00
5) Chlorobenzene-d5	8.863	117	7967	0.50	ug/L	# 0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	4009	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	3370	0.49	ug/L	0.00
Spiked Amount	0.500	Range 40 - 130	Recovery	=	98.00%	
4) 1,2-Dichloroethane-d4	5.042	65	3432	0.51	ug/L	0.00
Spiked Amount	0.500	Range 70 - 130	Recovery	=	102.00%	
7) 1,2-Dichloropropane-d6	6.068	67	4145	0.51	ug/L	0.00
Spiked Amount	0.500	Range 60 - 140	Recovery	=	102.00%	
8) Toluene-d8	7.317	98	9106	0.49	ug/L	0.00
Spiked Amount	0.500	Range 70 - 130	Recovery	=	98.00%	
10) 1,1,2,2-Tetrachloroeth...	10.216	84	2367	0.49	ug/L	0.00
Spiked Amount	0.500	Range 65 - 120	Recovery	=	98.00%	
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	6044	0.50	ug/L	99
6) Trichloroethene	5.925	95	4060	0.49	ug/L	100
9) 1,2-Dibromoethane	8.368	107	2232	0.51	ug/L	99
12) 1,2,3-Trichloropropane	10.267	75	2223	0.51	ug/L #	70
13) 1,2-Dibromo-3-chloropr...	12.426	75	284	0.46	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV010621\
Data File : VV019907.D
Acq On : 06 Jan 2021 12:35
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV205

Quant Time: Jan 07 06:48:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM010621.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jan 07 06:31:40 2021
Response via : Initial Calibration

