

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV010621\
 Data File : VV019910.D
 Acq On : 06 Jan 2021 14:19
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5281

Quant Time: Jan 07 06:36:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM010621.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jan 07 06:31:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	8288	0.50	ug/L	# 0.00
5) Chlorobenzene-d5	8.864	117	7830	0.50	ug/L	# 0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	3895	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	3393	0.49	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	98.00%		
4) 1,2-Dichloroethane-d4	5.042	65	3379	0.49	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	98.00%		
7) 1,2-Dichloropropane-d6	6.069	67	4136	0.52	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	104.00%		
8) Toluene-d8	7.317	98	8914	0.49	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	98.00%		
10) 1,1,2,2-Tetrachloroeth...	10.216	84	2297	0.49	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	98.00%		
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	6290	0.51	ug/L	98
6) Trichloroethene	5.925	95	4136	0.51	ug/L	99
9) 1,2-Dibromoethane	8.369	107	2238	0.52	ug/L	99
12) 1,2,3-Trichloropropane	10.267	75	2189	0.51	ug/L #	69
13) 1,2-Dibromo-3-chloropr...	12.427	75	284	0.48	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV010621\
Data File : VV019910.D
Acq On : 06 Jan 2021 14:19
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5281

Quant Time: Jan 07 06:36:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM010621.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jan 07 06:31:40 2021
Response via : Initial Calibration

