

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW010622\
 Data File : VW024296.D
 Acq On : 06 Jan 2022 15:27
 Operator : SY/MD
 Sample : N1053-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGLH5

Quant Time: Jan 07 01:38:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 07 01:34:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	165952	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	157040	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	76793	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	67662	56.652	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	113.300%	
7) Chloroethane-d5	1.571	69	55426	61.090	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	122.180%	
11) 1,1-Dichloroethene-d2	2.111	63	104571	46.847	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	93.700%	
21) 2-Butanone-d5	3.886	46	80961	121.670	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	121.670%	
24) Chloroform-d	4.349	84	139468	63.893	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	127.780%#	
26) 1,2-Dichloroethane-d4	5.031	65	94614	64.739	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	129.480%#	
32) Benzene-d6	5.050	84	262660	64.094	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	128.180%#	
36) 1,2-Dichloropropane-d6	6.069	67	76090	66.252	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	132.500%#	
41) Toluene-d8	7.317	98	249183	62.801	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	125.600%#	
43) trans-1,3-Dichloroprop...	7.622	79	39870	60.287	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	120.580%	
47) 2-Hexanone-d5	8.088	63	57537	115.310	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	=	115.310%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	98438	59.976	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	119.960%	
66) 1,2-Dichlorobenzene-d4	11.622	152	107000	69.904	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	139.800%#	
Target Compounds						
5) Vinyl chloride	1.314	62	10967	9.215	ug/L	91
20) cis-1,2-Dichloroethene	3.915	96	5053	4.542	ug/L	91
34) Trichloroethene	5.918	95	13382	12.153	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV010622\
Data File : VV024296.D
Acq On : 06 Jan 2022 15:27
Operator : SY/MD
Sample : N1053-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGLH5

Quant Time: Jan 07 01:38:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 07 01:34:41 2022
Response via : Initial Calibration

