

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV010721\
 Data File : VV019917.D
 Acq On : 07 Jan 2021 13:02
 Operator : SY/MD
 Sample : VIBLK72
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK72

Quant Time: Jan 08 01:55:23 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 08 01:32:18 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	537073	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	511447	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.26	152	243821	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	180234	43.37	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.74%
7) Chloroethane-d5	1.57	69	157097	47.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.10%
11) 1,1-Dichloroethene-d2	2.11	63	279254	35.55	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.10%
21) 2-Butanone-d5	3.89	46	289835	94.09	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.09%
24) Chloroform-d	4.35	84	361025	45.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.96%
26) 1,2-Dichloroethane-d4	5.04	65	255960	49.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.74%
32) Benzene-d6	5.05	84	702605	47.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.42%
36) 1,2-Dichloropropane-d6	6.07	67	229070	47.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.60%
41) Toluene-d8	7.32	98	601061	45.85	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.70%
43) trans-1,3-Dichloropropene-	7.63	79	101909	42.33	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.66%
47) 2-Hexanone-d5	8.09	63	163550	81.72	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.72%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	274804	43.63	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.26%
64) 1,2-Dichlorobenzene-d4	11.63	152	223344	47.89	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.78%

Target Compounds					Ovalue
46) Tetrachloroethene	7.98	164	4129	1.449 ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV010721\
Data File : VV019917.D
Acq On : 07 Jan 2021 13:02
Operator : SY/MD
Sample : VIBLK72
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK72

Quant Time: Jan 08 01:55:23 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 08 01:32:18 2021
Response via : Initial Calibration

