

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV010721\
 Data File : VV019914.D
 Acq On : 07 Jan 2021 11:52
 Operator : SY/MD
 Sample : M1017-07 100X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG4W7

Quant Time: Jan 08 01:45:46 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 08 01:32:18 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	505791	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	483929	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.25	152	225398	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	126908	32.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	64.84%
7) Chloroethane-d5	1.57	69	123100	39.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.14%
11) 1,1-Dichloroethene-d2	2.11	63	213920	28.92	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	57.84%#
21) 2-Butanone-d5	3.89	46	283005	97.55	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.55%
24) Chloroform-d	4.35	84	306271	41.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.84%
26) 1,2-Dichloroethane-d4	5.03	65	211915	43.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.68%
32) Benzene-d6	5.05	84	552021	39.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.22%
36) 1,2-Dichloropropane-d6	6.07	67	194619	42.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	84.94%
41) Toluene-d8	7.32	98	471279	38.00	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	76.00%#
43) trans-1,3-Dichloropropene-	7.62	79	82242	36.11	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.22%
47) 2-Hexanone-d5	8.09	63	155796	82.27	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.27%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	256558	43.05	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.10%
64) 1,2-Dichlorobenzene-d4	11.63	152	198573	46.06	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.12%


Target Compounds					Ovalue
34) Trichloroethene	5.92	95	17711	4.386 ug/L	98
46) Tetrachloroethene	7.98	164	269834	100.063 ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV010721\
Data File : VV019914.D
Acq On : 07 Jan 2021 11:52
Operator : SY/MD
Sample : M1017-07 100X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG4W7

Quant Time: Jan 08 01:45:46 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 08 01:32:18 2021
Response via : Initial Calibration









#34
 Trichloroethene
 Concen: 4.386 ug/L
 RT: 5.92 min Scan# 1502
 Delta R.T. 0.00 min
 Lab File: VV019914.D
 Acq: 07 Jan 2021 11:52

Instrument :
 MSVOA_V
 ClientSampleId :
 BG4W7

Tot Ion:	95	Resp:	17711
Ion	Ratio	Lower	Upper
95	100		
97	66.7	44.5	82.7
132	91.3	63.4	117.8
130	98.5	67.1	124.5



#46
 Tetrachloroethene
 Concen: 100.063 ug/L
 RT: 7.98 min Scan# 2142
 Delta R.T. 0.00 min
 Lab File: VV019914.D
 Acq: 07 Jan 2021 11:52

Tgt Ion:	164	Resp:	269834
Ion	Ratio	Lower	Upper
164	100		
129	104.3	70.6	131.0
131	99.1	66.8	124.0
166	127.3	88.2	163.8

