

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV010819\
 Data File : VV009198.D
 Acq On : 08 Jan 2019 17:28
 Operator : SY/MD
 Sample : VV0108SBL01
 Misc : 5.00 G/10ML/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK78

Quant Time: Jan 09 00:46:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM010819S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 09 00:40:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	251768	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	240136	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	119500	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	61316	26.59	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	106.36%
7) Chloroethane-d5	1.58	69	47041	23.51	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	94.04%
10) 1,1-Dichloroethene-d2	2.13	63	123442	19.60	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	78.40%
20) 2-Butanone-d5	3.96	46	47583	47.63	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.26%
24) Chloroform-d	4.40	84	145076	24.76	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	99.04%
26) 1,2-Dichloroethane-d4	5.08	65	83999	26.36	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	105.44%
29) Benzene-d6	5.10	84	303928	25.24	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	100.96%
33) 1,2-Dichloropropane-d6	6.12	67	87930	25.51	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	102.04%
37) Toluene-d8	7.36	98	284432	24.97	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	99.88%
38) trans-1,3-Dichloropropene-	7.67	79	41998	24.79	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	99.16%
39) 2-Hexanone-d5	8.14	63	30560	48.53	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.06%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	85397	25.46	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	101.84%
61) 1,2-Dichlorobenzene-d4	11.68	152	112006	24.94	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.76%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV010819\
Data File : VV009198.D
Acq On : 08 Jan 2019 17:28
Operator : SY/MD
Sample : VV0108SBL01
Misc : 5.00 G/10ML/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK78

Quant Time: Jan 09 00:46:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM010819S.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 09 00:40:20 2019
Response via : Initial Calibration

