

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV010821\
 Data File : VV019948.D
 Acq On : 08 Jan 2021 15:22
 Operator : SY/MD
 Sample : VIBLK77
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK77

Quant Time: Jan 09 00:07:46 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 08 22:51:06 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	444873	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	423035	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.26	152	191896	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	124847	36.26	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	72.52%
7) Chloroethane-d5	1.57	69	129375	47.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.56%
11) 1,1-Dichloroethene-d2	2.11	63	200509	30.81	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.62%
21) 2-Butanone-d5	3.89	46	287669	112.74	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	112.74%
24) Chloroform-d	4.35	84	319367	49.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.20%
26) 1,2-Dichloroethane-d4	5.04	65	237776	55.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.86%
32) Benzene-d6	5.05	84	585460	48.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.12%
36) 1,2-Dichloropropane-d6	6.07	67	204817	51.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.26%
41) Toluene-d8	7.32	98	472216	43.55	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.10%
43) trans-1,3-Dichloropropene-	7.63	79	86821	43.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.20%
47) 2-Hexanone-d5	8.09	63	158384	95.67	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.67%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	257364	49.40	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.80%
64) 1,2-Dichlorobenzene-d4	11.63	152	183705	50.05	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.10%

Target Compounds					Ovalue
46) Tetrachloroethene	7.98	164	3982	1.689 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV010821\
Data File : VV019948.D
Acq On : 08 Jan 2021 15:22
Operator : SY/MD
Sample : VIBLK77
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK77

Quant Time: Jan 09 00:07:46 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 08 22:51:06 2021
Response via : Initial Calibration

