

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV010821\
 Data File : VV019952.D
 Acq On : 08 Jan 2021 17:11
 Operator : SY/MD
 Sample : VIBLK78
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK78

Quant Time: Jan 09 00:34:56 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 08 22:51:06 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	490697	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	468090	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.26	152	221789	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	146205	38.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.00%
7) Chloroethane-d5	1.57	69	137250	45.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.94%
11) 1,1-Dichloroethene-d2	2.11	63	238714	33.26	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.52%
21) 2-Butanone-d5	3.89	46	279253	99.22	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.22%
24) Chloroform-d	4.35	84	323150	45.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.08%
26) 1,2-Dichloroethane-d4	5.04	65	229266	48.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.78%
32) Benzene-d6	5.05	84	611239	45.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.70%
36) 1,2-Dichloropropane-d6	6.07	67	205412	46.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.70%
41) Toluene-d8	7.32	98	519625	43.31	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.62%
43) trans-1,3-Dichloropropene-	7.63	79	88167	40.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.04%
47) 2-Hexanone-d5	8.09	63	153929	84.03	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.03%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	256801	44.55	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.10%
64) 1,2-Dichlorobenzene-d4	11.63	152	202615	47.76	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.52%

Target Compounds					Ovalue
46) Tetrachloroethene	7.98	164	1916	0.735 ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV010821\
Data File : VV019952.D
Acq On : 08 Jan 2021 17:11
Operator : SY/MD
Sample : VIBLK78
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK78

Quant Time: Jan 09 00:34:56 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 08 22:51:06 2021
Response via : Initial Calibration

