

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV010821\
 Data File : VV019942.D
 Acq On : 08 Jan 2021 12:59
 Operator : SY/MD
 Sample : M1017-04
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jan 08 23:38:23 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 08 22:51:06 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	466906	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	453457	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.26	152	212126	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	136276	37.72	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.44%
7) Chloroethane-d5	1.57	69	132280	46.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.12%
11) 1,1-Dichloroethene-d2	2.11	63	219752	32.18	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.36%
21) 2-Butanone-d5	3.89	46	278712	104.07	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.07%
24) Chloroform-d	4.35	84	325221	47.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.28%
26) 1,2-Dichloroethane-d4	5.04	65	239176	53.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.20%
32) Benzene-d6	5.05	84	605959	46.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.82%
36) 1,2-Dichloropropane-d6	6.07	67	208237	48.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.00%
41) Toluene-d8	7.32	98	507271	43.65	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.30%
43) trans-1,3-Dichloropropene-	7.63	79	89200	41.79	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.58%
47) 2-Hexanone-d5	8.09	63	153213	86.34	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.34%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	257261	46.07	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.14%
64) 1,2-Dichlorobenzene-d4	11.63	152	200635	49.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.90%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV010821\
Data File : VV019942.D
Acq On : 08 Jan 2021 12:59
Operator : SY/MD
Sample : M1017-04
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jan 08 23:38:23 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM122820WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 08 22:51:06 2021
Response via : Initial Calibration

