

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV010822\
 Data File : VV024310.D
 Acq On : 08 Jan 2022 11:32
 Operator : SY/MD
 Sample : VV0108WBL03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK291

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/10/2022
 Supervised By : Mahesh Dadoda 01/10/2022

Quant Time: Jan 10 02:05:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR010422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jan 10 01:53:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	101710	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	100115	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	54206	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	32054	3.880	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.600%
7) Chloroethane-d5	1.568	69	28101	4.328	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.600%
11) 1,1-Dichloroethene-d2	2.108	63	49026	3.026	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.600%
20) 2-Butanone-d5	3.915	46	45178m	42.072	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.140%
24) Chloroform-d	4.342	84	66503	4.297	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.000%
26) 1,2-Dichloroethane-d4	5.031	65	33365	4.557	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.200%
32) Benzene-d6	5.047	84	125487	4.475	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.400%
36) 1,2-Dichloropropane-d6	6.066	67	33913	4.653	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.000%
41) Toluene-d8	7.313	98	110424	3.964	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.200%
43) trans-1,3-Dichloroprop...	7.622	79	13793	4.131	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.600%
46) 2-Hexanone-d5	8.088	63	43657	41.843	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.680%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	22282	4.201	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	45475	4.830	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV010822\
Data File : VV024310.D
Acq On : 08 Jan 2022 11:32
Operator : SY/MD
Sample : VV0108WBL03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK291

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/10/2022
Supervised By : Mahesh Dadoda 01/10/2022

Quant Time: Jan 10 02:05:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR010422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jan 10 01:53:06 2022
Response via : Initial Calibration

