

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011024\
 Data File : VW033567.D
 Acq On : 10 Jan 2024 17:54
 Operator : SY/MD
 Sample : P1061-13
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCMH3

Quant Time: Jan 11 00:52:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 11 00:32:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	329824	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	305975	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	147498	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	121408	47.895	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	95.800%	
7) Chloroethane-d5	1.529	69	94608	47.129	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	94.260%	
11) 1,1-Dichloroethene-d2	2.060	65	51027	46.774	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	93.540%	
21) 2-Butanone-d5	3.783	46	163427	93.995	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	94.000%	
24) Chloroform-d	4.249	84	204738	45.170	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	90.340%	
26) 1,2-Dichloroethane-d4	4.941	65	139657	47.687	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	95.380%	
32) Benzene-d6	4.960	84	423822	48.723	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.440%	
36) 1,2-Dichloropropane-d6	5.989	67	142135	49.640	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	99.280%	
41) Toluene-d8	7.243	98	379782	48.457	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	96.920%	
43) trans-1,3-Dichloroprop...	7.555	79	68495	48.402	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	96.800%	
47) 2-Hexanone-d5	8.024	63	120532	101.484	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	101.480%	
56) 1,1,2,2-Tetrachloroeth...	10.153	84	173828	47.554	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	95.100%	
66) 1,2-Dichlorobenzene-d4	11.561	152	137028	48.966	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	97.940%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011024\
Data File : VV033567.D
Acq On : 10 Jan 2024 17:54
Operator : SY/MD
Sample : P1061-13
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCMH3

Quant Time: Jan 11 00:52:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 11 00:32:34 2024
Response via : Initial Calibration

