

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011024\
 Data File : VW033574.D
 Acq On : 10 Jan 2024 20:41
 Operator : SY/MD
 Sample : P1059-05
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Jan 11 00:55:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 11 00:32:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	326720	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	302424	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	146060	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	125546	49.998	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	100.000%
7) Chloroethane-d5	1.529	69	97535	49.049	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.100%
11) 1,1-Dichloroethene-d2	2.060	65	53014	49.057	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.120%
21) 2-Butanone-d5	3.783	46	164698	95.626	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.630%
24) Chloroform-d	4.249	84	208930	46.533	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.060%
26) 1,2-Dichloroethane-d4	4.941	65	143659	49.519	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.040%
32) Benzene-d6	4.960	84	437328	50.866	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.740%
36) 1,2-Dichloropropane-d6	5.989	67	143026	50.538	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.080%
41) Toluene-d8	7.243	98	385831	49.807	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.620%
43) trans-1,3-Dichloroprop...	7.555	79	68201	48.761	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.520%
47) 2-Hexanone-d5	8.021	63	120818	102.919	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.920%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	175964	48.703	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.400%
66) 1,2-Dichlorobenzene-d4	11.561	152	140005	50.522	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011024\
Data File : VV033574.D
Acq On : 10 Jan 2024 20:41
Operator : SY/MD
Sample : P1059-05
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Jan 11 00:55:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 11 00:32:34 2024
Response via : Initial Calibration

