

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011121\
 Data File : VV019968.D
 Acq On : 11 Jan 2021 15:54
 Operator : SY/MD
 Sample : VV0111MBL01
 Misc : 5.0g/5.0ml/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK72

Quant Time: Jan 12 06:41:20 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM011121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Jan 11 13:51:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.621	114	275698	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.859	117	263539	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.257	152	129200	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.309	65	174384	74.99	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 149.98%#			
7) Chloroethane-d5	1.570	69	142583	76.81	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 153.62%#			
11) 1,1-Dichloroethene-d2	2.110	63	256174	55.93	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 111.86%			
21) 2-Butanone-d5	3.891	46	213028	141.21	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 141.21%#			
24) Chloroform-d	4.354	84	293385	70.04	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 140.08%#			
26) 1,2-Dichloroethane-d4	5.036	65	212492	74.62	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 149.24%#			
32) Benzene-d6	5.055	84	554494	74.55	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 149.10%#			
36) 1,2-Dichloropropane-d6	6.074	67	175469	75.56	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 151.12%#			
41) Toluene-d8	7.322	98	484596	71.89	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 143.78%#			
43) trans-1,3-Dichloroprop...	7.624	79	81885	67.95	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 135.90%#			
47) 2-Hexanone-d5	8.093	63	121266	130.25	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 130.25%#			
57) 1,1,2,2-Tetrachloroeth...	10.222	84	211447	67.35	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 134.70%#			
64) 1,2-Dichlorobenzene-d4	11.630	152	186654	74.80	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 149.60%#			

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011121\
Data File : VV019968.D
Acq On : 11 Jan 2021 15:54
Operator : SY/MD
Sample : VV0111MBL01
Misc : 5.0g/5.0ml/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK72

Quant Time: Jan 12 06:41:20 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM011121WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Jan 11 13:51:52 2021
Response via : Initial Calibration

