

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011121\
 Data File : VW019970.D
 Acq On : 11 Jan 2021 16:41
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK

Quant Time: Jan 12 06:41:58 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM011121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Jan 11 13:51:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.621	114	403361	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.858	117	382823	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.257	152	192317	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.309	65	162721	47.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.66%
7) Chloroethane-d5	1.569	69	133672	49.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.44%
11) 1,1-Dichloroethene-d2	2.109	63	248536	37.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.18%
21) 2-Butanone-d5	3.900	46	218137	98.83	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.83%
24) Chloroform-d	4.354	84	292173	47.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.34%
26) 1,2-Dichloroethane-d4	5.039	65	209996	50.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.80%
32) Benzene-d6	5.055	84	547384	50.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.32%
36) 1,2-Dichloropropane-d6	6.074	67	172114	51.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.04%
41) Toluene-d8	7.321	98	483291	49.35	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.70%
43) trans-1,3-Dichloroprop...	7.627	79	80543	46.01	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.02%
47) 2-Hexanone-d5	8.093	63	121427	89.78	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.78%
57) 1,1,2,2-Tetrachloroeth...	10.222	84	211448	46.37	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.74%
64) 1,2-Dichlorobenzene-d4	11.633	152	189821	51.10	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.20%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011121\
Data File : VV019970.D
Acq On : 11 Jan 2021 16:41
Operator : SY/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK

Quant Time: Jan 12 06:41:58 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM011121WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Jan 11 13:51:52 2021
Response via : Initial Calibration

