

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011121\
 Data File : VV019972.D
 Acq On : 11 Jan 2021 17:28
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK

Quant Time: Jan 12 06:42:29 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM011121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Jan 11 13:51:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.624	114	368263	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.858	117	357874	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.257	152	175850	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.309	65	159454	51.33	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 102.66%			
7) Chloroethane-d5	1.569	69	133715	53.93	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 107.86%			
11) 1,1-Dichloroethene-d2	2.109	63	240282	39.27	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 78.54%			
21) 2-Butanone-d5	3.894	46	208628	103.53	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 103.53%			
24) Chloroform-d	4.354	84	282427	50.48	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 100.96%			
26) 1,2-Dichloroethane-d4	5.035	65	204768	53.83	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 107.66%			
32) Benzene-d6	5.055	84	521031	51.58	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 103.16%			
36) 1,2-Dichloropropane-d6	6.074	67	167201	53.02	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 106.04%			
41) Toluene-d8	7.321	98	459680	50.21	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 100.42%			
43) trans-1,3-Dichloroprop...	7.627	79	77928	47.62	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 95.24%			
47) 2-Hexanone-d5	8.093	63	118105	93.42	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 93.42%			
57) 1,1,2,2-Tetrachloroeth...	10.222	84	204295	47.92	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 95.84%			
64) 1,2-Dichlorobenzene-d4	11.630	152	186105	54.79	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 109.58%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011121\
Data File : VV019972.D
Acq On : 11 Jan 2021 17:28
Operator : SY/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK

Quant Time: Jan 12 06:42:29 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM011121WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Jan 11 13:51:52 2021
Response via : Initial Calibration

