

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011224\
 Data File : VW033603.D
 Acq On : 12 Jan 2024 18:19
 Operator : SY/MD
 Sample : P1108-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B37

Quant Time: Jan 12 23:47:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 12 23:30:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	129059	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	121552	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	60120	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	32504	4.606	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.200%
7) Chloroethane-d5	1.535	69	29509	4.715	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.200%
11) 1,1-Dichloroethene-d2	2.060	65	14936	4.696	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	94.000%
20) 2-Butanone-d5	3.809	46	87213	50.544	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.080%
24) Chloroform-d	4.252	84	72682	4.762	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.200%
26) 1,2-Dichloroethane-d4	4.947	65	35791	5.230	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.600%
32) Benzene-d6	4.963	84	142304	4.852	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.000%
36) 1,2-Dichloropropane-d6	5.992	67	47013	4.808	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.200%
41) Toluene-d8	7.246	98	126255	4.822	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.400%
43) trans-1,3-Dichloroprop...	7.561	79	15444	4.569	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.400%
46) 2-Hexanone-d5	8.027	63	65883	49.639	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.280%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	28631	5.011	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.200%
66) 1,2-Dichlorobenzene-d4	11.561	152	46311	5.017	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011224\
Data File : VV033603.D
Acq On : 12 Jan 2024 18:19
Operator : SY/MD
Sample : P1108-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B37

Quant Time: Jan 12 23:47:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 12 23:30:52 2024
Response via : Initial Calibration

