

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011224\
 Data File : VW033606.D
 Acq On : 12 Jan 2024 19:30
 Operator : SY/MD
 Sample : P1108-16
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B55

Quant Time: Jan 12 23:49:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 12 23:30:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	126269	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	117429	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	59723	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	31781	4.603	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.000%
7) Chloroethane-d5	1.535	69	28600	4.670	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.400%
11) 1,1-Dichloroethene-d2	2.063	65	14776	4.749	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	95.000%
20) 2-Butanone-d5	3.805	46	83384	49.392	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.780%
24) Chloroform-d	4.252	84	71174	4.767	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.400%
26) 1,2-Dichloroethane-d4	4.947	65	34556	5.161	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.200%
32) Benzene-d6	4.963	84	140346	4.953	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.000%
36) 1,2-Dichloropropane-d6	5.992	67	44614	4.723	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.400%
41) Toluene-d8	7.246	98	122138	4.828	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.600%
43) trans-1,3-Dichloroprop...	7.564	79	15797	4.837	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.800%
46) 2-Hexanone-d5	8.027	63	64688	50.450	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.900%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	28184	5.106	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.200%
66) 1,2-Dichlorobenzene-d4	11.561	152	45027	4.910	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011224\
Data File : VV033606.D
Acq On : 12 Jan 2024 19:30
Operator : SY/MD
Sample : P1108-16
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B55

Quant Time: Jan 12 23:49:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 12 23:30:52 2024
Response via : Initial Calibration

