

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011224\
 Data File : VW033609.D
 Acq On : 12 Jan 2024 21:06
 Operator : SY/MD
 Sample : P1108-17
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B57

Quant Time: Jan 13 00:30:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jan 13 00:29:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	124199	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	113381	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	58022	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	29887	4.401	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.000%
7) Chloroethane-d5	1.532	69	28588	4.746	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.000%
11) 1,1-Dichloroethene-d2	2.060	65	14153	4.624	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.400%
20) 2-Butanone-d5	3.831	46	88619	53.368	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.740%
24) Chloroform-d	4.252	84	66463	4.525	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.600%
26) 1,2-Dichloroethane-d4	4.947	65	32762	4.975	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.600%
32) Benzene-d6	4.963	84	131286	4.799	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.000%
36) 1,2-Dichloropropane-d6	5.992	67	43778	4.800	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.000%
41) Toluene-d8	7.246	98	113901	4.664	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.200%
43) trans-1,3-Dichloroprop...	7.561	79	15578	4.940	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.800%
46) 2-Hexanone-d5	8.030	63	64018	51.710	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.420%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	28034	5.260	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.200%
66) 1,2-Dichlorobenzene-d4	11.561	152	43241	4.854	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011224\
Data File : VV033609.D
Acq On : 12 Jan 2024 21:06
Operator : SY/MD
Sample : P1108-17
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B57

Quant Time: Jan 13 00:30:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 13 00:29:44 2024
Response via : Initial Calibration

