

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011224\
 Data File : VW033620.D
 Acq On : 13 Jan 2024 01:27
 Operator : SY/MD
 Sample : P1121-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH8L6

Quant Time: Jan 13 03:59:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jan 13 00:29:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	115581	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	109051	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	54930	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	29960	4.740	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	94.800%	
7) Chloroethane-d5	1.535	69	27481	4.903	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	98.000%	
11) 1,1-Dichloroethene-d2	2.063	65	13590	4.771	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	95.400%	
20) 2-Butanone-d5	3.799	46	77448	50.118	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	100.240%	
24) Chloroform-d	4.252	84	66500	4.865	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.400%	
26) 1,2-Dichloroethane-d4	4.947	65	33402	5.450	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	109.000%	
32) Benzene-d6	4.963	84	129621	4.926	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.600%	
36) 1,2-Dichloropropane-d6	5.992	67	43067	4.909	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	98.200%	
41) Toluene-d8	7.246	98	113431	4.829	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	96.600%	
43) trans-1,3-Dichloroprop...	7.564	79	13750	4.534	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	90.600%	
46) 2-Hexanone-d5	8.027	63	63718	53.512	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	107.020%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	27216	5.310	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	106.200%	
66) 1,2-Dichlorobenzene-d4	11.564	152	43596	5.169	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	103.400%	
Target Compounds						Qvalue
13) Acetone	2.124	43	8759	6.982	ug/L	99
14) Carbon disulfide	2.243	76	36691	1.217	ug/L	98
22) cis-1,2-Dichloroethene	3.825	96	2078	0.216	ug/L	83
34) Trichloroethene	5.847	95	3054	0.296	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011224\
Data File : VV033620.D
Acq On : 13 Jan 2024 01:27
Operator : SY/MD
Sample : P1121-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH8L6

Quant Time: Jan 13 03:59:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 13 00:29:44 2024
Response via : Initial Calibration

