

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011224\
 Data File : VW033624.D
 Acq On : 13 Jan 2024 03:03
 Operator : SY/MD
 Sample : P1121-17
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH8M0

Quant Time: Jan 13 04:01:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jan 13 00:29:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	117208	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	107902	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	54871	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	31519	4.918	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.400%
7) Chloroethane-d5	1.532	69	28598	5.031	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.600%
11) 1,1-Dichloroethene-d2	2.060	65	14403	4.987	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	99.800%
20) 2-Butanone-d5	3.815	46	86777	55.376	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.760%
24) Chloroform-d	4.253	84	71249	5.141	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.800%
26) 1,2-Dichloroethane-d4	4.947	65	35420	5.700	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.000%
32) Benzene-d6	4.963	84	133912	5.143	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.800%
36) 1,2-Dichloropropane-d6	5.992	67	44590	5.137	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.800%
41) Toluene-d8	7.246	98	119821	5.155	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.200%
43) trans-1,3-Dichloroprop...	7.561	79	14045	4.680	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.600%
46) 2-Hexanone-d5	8.031	63	68388	58.045	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	116.100%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	28918	5.702	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	114.000%
66) 1,2-Dichlorobenzene-d4	11.561	152	44458	5.277	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011224\
Data File : VV033624.D
Acq On : 13 Jan 2024 03:03
Operator : SY/MD
Sample : P1121-17
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH8M0

Quant Time: Jan 13 04:01:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 13 00:29:44 2024
Response via : Initial Calibration

