

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011323\
 Data File : VW029853.D
 Acq On : 14 Jan 2023 12:53
 Operator : SY/MD
 Sample : 01149-06
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0L56

Quant Time: Jan 16 00:40:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 12 22:05:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	210785	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	211259	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	110120	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	66804	4.178	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.600%
7) Chloroethane-d5	1.564	69	63034	4.756	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.200%
11) 1,1-Dichloroethene-d2	2.104	63	94902	3.256	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.200%
20) 2-Butanone-d5	3.895	46	143608	50.678	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.360%
24) Chloroform-d	4.339	84	143632	4.897	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.000%
26) 1,2-Dichloroethane-d4	5.024	65	64523	5.195	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.800%
32) Benzene-d6	5.043	84	293297	4.680	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.600%
36) 1,2-Dichloropropane-d6	6.063	67	92185	5.026	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.600%
41) Toluene-d8	7.307	98	253625	4.372	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.400%
43) trans-1,3-Dichloroprop...	7.616	79	28049	4.188	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.800%
46) 2-Hexanone-d5	8.085	63	127938	46.917	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.840%
56) 1,1,2,2-Tetrachloroeth...	10.204	84	66722	4.964	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.200%
66) 1,2-Dichlorobenzene-d4	11.612	152	99424	4.966	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.400%
Target Compounds						
17) Methyl tert-butyl Ether	2.767	73	31346	1.007	ug/L	97
22) cis-1,2-Dichloroethene	3.908	96	6487	0.364	ug/L	91
34) Trichloroethene	5.905	95	293611	16.587	ug/L	97
42) Toluene	7.381	91	98792	1.311	ug/L	99
53) m,p-Xylene	9.127	106	1808	0.058	ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011323\
Data File : VV029853.D
Acq On : 14 Jan 2023 12:53
Operator : SY/MD
Sample : 01149-06
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 60 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COL56

Quant Time: Jan 16 00:40:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 12 22:05:54 2023
Response via : Initial Calibration

