

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011323\
 Data File : VW029816.D
 Acq On : 13 Jan 2023 18:17
 Operator : SY/MD
 Sample : 01138-10
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH1X3

Quant Time: Jan 13 22:52:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 12 22:05:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	199549	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	196585	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	102104	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	72271	4.775	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.400%
7) Chloroethane-d5	1.565	69	63953	5.097	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.000%
11) 1,1-Dichloroethene-d2	2.105	63	99131	3.593	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.800%
20) 2-Butanone-d5	3.886	46	106944	39.865	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.720%
24) Chloroform-d	4.339	84	143061	5.153	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.000%
26) 1,2-Dichloroethane-d4	5.024	65	62619	5.325	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.600%
32) Benzene-d6	5.043	84	300000	5.145	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.800%
36) 1,2-Dichloropropane-d6	6.063	67	92396	5.414	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.200%
41) Toluene-d8	7.310	98	263120	4.874	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.400%
43) trans-1,3-Dichloroprop...	7.616	79	25617	4.110	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.200%
46) 2-Hexanone-d5	8.085	63	100656	39.668	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.340%
56) 1,1,2,2-Tetrachloroeth...	10.204	84	63465	5.075	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.400%
66) 1,2-Dichlorobenzene-d4	11.612	152	102633	5.529	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.600%
Target Compounds						
					Qvalue	
3) Chloromethane	1.240	50	2090	0.107	ug/L	97
5) Vinyl chloride	1.307	62	31798	1.659	ug/L	92
8) Chloroethane	1.581	64	48980	4.212	ug/L #	70
12) 1,1-Dichloroethene	2.114	96	8588	0.504	ug/L #	1
14) Carbon disulfide	2.294	76	6334	0.129	ug/L	99
18) trans-1,2-Dichloroethene	2.757	96	26814	1.693	ug/L	97
19) 1,1-Dichloroethane	3.185	63	44138	1.666	ug/L	98
22) cis-1,2-Dichloroethene	3.905	96	9743	0.577	ug/L	95
34) Trichloroethene	5.908	95	15214	0.924	ug/L	95
53) m,p-Xylene	9.133	106	2514	0.087	ug/L	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011323\
Data File : VV029816.D
Acq On : 13 Jan 2023 18:17
Operator : SY/MD
Sample : 01138-10
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH1X3

Quant Time: Jan 13 22:52:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 12 22:05:54 2023
Response via : Initial Calibration









#3
Chloromethane
Concen: 0.107 ug/L
RT: 1.240 min Scan# 61
Delta R.T. 0.003 min
Lab File: VV029816.D
Acq: 13 Jan 2023 18:17




Instrument :
MSVOA_V
ClientSampleId :
BH1X3


Tgt Ion: 50 Resp: 2090
Ion Ratio Lower Upper
50 100
52 31.4 23.1 42.9



#5
Vinyl chloride
Concen: 1.659 ug/L
RT: 1.307 min Scan# 83
Delta R.T. 0.003 min
Lab File: VV029816.D
Acq: 13 Jan 2023 18:17


Tgt Ion: 62 Resp: 31798
Ion Ratio Lower Upper
62 100
64 37.9 23.2 43.2








#8
Chloroethane
Concen: 4.212 ug/L
RT: 1.581 min Scan# 1
Delta R.T. 0.003 min
Lab File: VV029816.D
Acq: 13 Jan 2023 18:17


Instrument :
MSVOA_V
ClientSampleId :
BH1X3





Tgt Ion: 64 Resp: 48980
Ion Ratio Lower Upper
64 100
66 16.1 22.9 42.5#




#12
1,1-Dichloroethene
Concen: 0.504 ug/L
RT: 2.114 min Scan# 334
Delta R.T. 0.003 min
Lab File: VV029816.D
Acq: 13 Jan 2023 18:17



Tgt Ion: 96 Resp: 8588
Ion Ratio Lower Upper
96 100
61 209.8 106.1 197.1#
63 730.4 76.4 141.8#









#14
Carbon disulfide
Concen: 0.129 ug/L
RT: 2.294 min Scan# 390
Delta R.T. 0.010 min
Lab File: VV029816.D
Acq: 13 Jan 2023 18:17




Instrument :
MSVOA_V
ClientSampleId :
BH1X3


Tgt Ion: 76 Resp: 6334
Ion Ratio Lower Upper
76 100
78 8.9 7.4 11.0



#18
trans-1,2-Dichloroethene
Concen: 1.693 ug/L
RT: 2.757 min Scan# 534
Delta R.T. 0.006 min
Lab File: VV029816.D
Acq: 13 Jan 2023 18:17

Tgt Ion: 96 Resp: 26814
Ion Ratio Lower Upper
96 100
61 128.3 86.1 159.9
98 63.0 43.9 81.5





#19

1,1-Dichloroethane

Concen: 1.666 ug/L

RT: 3.185 min Scan# 6

Delta R.T. 0.006 min

Lab File: VV029816.D

Acq: 13 Jan 2023 18:17

Instrument :

MSVOA_V

ClientSampleId :

BH1X3





Tgt Ion: 63 Resp: 44138

Ion Ratio Lower Upper

63 100

65 33.2 22.4 41.6

83 13.5 10.2 19.0



#22

cis-1,2-Dichloroethene

Concen: 0.577 ug/L

RT: 3.905 min Scan# 891

Delta R.T. 0.006 min

Lab File: VV029816.D

Acq: 13 Jan 2023 18:17




Tgt Ion: 96 Resp: 9743


Ion Ratio Lower Upper

96 100

61 120.1 80.0 148.6

68 0.0 0.0 0.0





#34

Trichloroethene

Concen: 0.924 ug/L

RT: 5.908 min Scan# 1

Delta R.T. 0.003 min

Lab File: VV029816.D

Acq: 13 Jan 2023 18:17

Instrument :

MSVOA_V

ClientSampleId :

BH1X3

Tgt Ion: 95 Resp: 15214





Ion Ratio Lower Upper

95 100

97 61.5 44.9 83.3

132 98.5 71.8 133.4

130 100.7 75.3 139.8



#53

m,p-Xylene

Concen: 0.087 ug/L

RT: 9.133 min Scan# 2517

Delta R.T. 0.003 min

Lab File: VV029816.D

Acq: 13 Jan 2023 18:17

Tgt Ion:106 Resp: 2514

Ion Ratio Lower Upper

106 100

91 186.6 138.7 257.7

