

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011323\
 Data File : VW029855.D
 Acq On : 14 Jan 2023 13:41
 Operator : SY/MD
 Sample : 01149-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Jan 16 00:40:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 12 22:05:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	178270	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	172471	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	82353	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	68911	5.096	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.000%
7) Chloroethane-d5	1.564	69	64721	5.774	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.400%
11) 1,1-Dichloroethene-d2	2.105	63	92285	3.744	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.800%
20) 2-Butanone-d5	3.896	46	141956	59.232	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.460%
24) Chloroform-d	4.339	84	140756	5.675	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.400%
26) 1,2-Dichloroethane-d4	5.024	65	62441	5.944	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	118.800%
32) Benzene-d6	5.043	84	277169	5.418	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.400%
36) 1,2-Dichloropropane-d6	6.063	67	86735	5.793	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	115.800%
41) Toluene-d8	7.307	98	237636	5.017	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.400%
43) trans-1,3-Dichloroprop...	7.616	79	27289	4.990	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.800%
46) 2-Hexanone-d5	8.085	63	126783	56.950	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	113.900%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	64336	5.863	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	117.200%
66) 1,2-Dichlorobenzene-d4	11.612	152	84447	5.640	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011323\
Data File : VV029855.D
Acq On : 14 Jan 2023 13:41
Operator : SY/MD
Sample : 01149-08
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 62 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Jan 16 00:40:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 12 22:05:54 2023
Response via : Initial Calibration

