

Quantitation Report (Qedit)




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011421\
 Data File : VV020000.D
 Acq On : 14 Jan 2021 14:36
 Operator : SY/MD
 Sample : M1147-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG4X6

Manual Integrations
 APPROVED

MMDadoda
 1/18/2021 12:42:57 PM

Quant Time: Jan 15 01:08:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM011121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 13 13:06:01 2021
 Response via : Initial Calibration



TIC: VV020000.D\data.ms

(13) Acetone (T)

2.177min (0.000) 3.44 ug/L

response 6020

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	30.10	34.97
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011421\
 Data File : VV020000.D
 Acq On : 14 Jan 2021 14:36
 Operator : SY/MD
 Sample : M1147-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG4X6

Manual Integrations
 APPROVED

MMDadoda
 1/18/2021 12:42:57 PM

Quant Time: Jan 15 01:08:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM011121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 13 13:06:01 2021
 Response via : Initial Calibration



TIC: VV020000.D\data.ms

(13) Acetone (T)

2.177min (0.000) 7.25 ug/L m

response 12701

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	30.10	16.57
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\W011421\
 Data File : W020000.D
 Acq On : 14 Jan 2021 14:36
 Operator : SY/MD
 Sample : M1147-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG4X6

Manual Integrations
 APPROVED

MMDadoda
 1/18/2021 12:42:57 PM

Quant Time: Jan 15 01:08:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM011121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 13 13:06:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.618	114	362382	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.855	117	351506	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.257	152	182772	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.309	65	154198	50.45	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	100.90%		
7) Chloroethane-d5	1.570	69	132335	54.24	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	108.48%		
11) 1,1-Dichloroethene-d2	2.110	63	235669	39.15	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	78.30%		
21) 2-Butanone-d5	3.885	46	185676	93.64	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	93.64%		
24) Chloroform-d	4.348	84	267850	48.65	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	97.30%		
26) 1,2-Dichloroethane-d4	5.032	65	191225	51.09	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	102.18%		
32) Benzene-d6	5.052	84	494802	49.88	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	99.76%		
36) 1,2-Dichloropropane-d6	6.071	67	156176	50.42	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	100.84%		
41) Toluene-d8	7.319	98	437758	48.69	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	97.38%		
43) trans-1,3-Dichloroprop...	7.624	79	71509	44.49	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	88.98%		
47) 2-Hexanone-d5	8.090	63	105833	85.23	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	85.23%		
57) 1,1,2,2-Tetrachloroeth...	10.222	84	190705	45.54	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	91.08%		
64) 1,2-Dichlorobenzene-d4	11.630	152	178752	50.64	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	101.28%		
Target Compounds						
13) Acetone	2.177	43	12701m	7.25	ug/L	Qvalue 85
48) 2-Hexanone	8.148	43	10049	3.14	ug/L	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011421\
Data File : VW020000.D
Acq On : 14 Jan 2021 14:36
Operator : SY/MD
Sample : M1147-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG4X6

Manual Integrations
APPROVED

MMDadoda
1/18/2021 12:42:57 PM

Quant Time: Jan 15 01:08:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM011121WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 13 13:06:01 2021
Response via : Initial Calibration

