

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011421\
 Data File : VV019997.D
 Acq On : 14 Jan 2021 13:12
 Operator : SY/MD
 Sample : M1147-03 100X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG4X6

Quant Time: Jan 15 03:14:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM011121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 13 13:06:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.621	114	389518	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.859	117	372707	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.254	152	188807	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.306	65	154693	47.08	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.16%
7) Chloroethane-d5	1.566	69	130764	49.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.72%
11) 1,1-Dichloroethene-d2	2.106	63	233279	36.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.10%
21) 2-Butanone-d5	3.891	46	203837	95.63	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.63%
24) Chloroform-d	4.354	84	272534	46.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.10%
26) 1,2-Dichloroethane-d4	5.036	65	195468	48.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.16%
32) Benzene-d6	5.055	84	505824	48.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.18%
36) 1,2-Dichloropropane-d6	6.074	67	158352	48.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.44%
41) Toluene-d8	7.318	98	443190	46.49	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.98%
43) trans-1,3-Dichloroprop...	7.627	79	71984	42.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.48%
47) 2-Hexanone-d5	8.093	63	111285	84.52	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.52%
57) 1,1,2,2-Tetrachloroeth...	10.222	84	195566	44.05	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.10%
64) 1,2-Dichlorobenzene-d4	11.633	152	177909	48.79	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.58%
Target Compounds						
48) 2-Hexanone	8.148	43	12198	3.60	ug/L	Qvalue # 82


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011421\
Data File : VV019997.D
Acq On : 14 Jan 2021 13:12
Operator : SY/MD
Sample : M1147-03 100X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG4X6


Quant Time: Jan 15 03:14:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM011121WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 13 13:06:01 2021
Response via : Initial Calibration





#48
 2-Hexanone
 Concen: 3.60 ug/L
 RT: 8.148 min Scan# 2195
 Delta R.T. 0.006 min
 Lab File: VV019997.D
 Acq: 14 Jan 2021 13:12

Instrument :
 MSVOA_V
 ClientSampleId :
 BG4X6



Tgt Ion: 43 Resp: 12198

Ion	Ratio	Lower	Upper
43	100		
58	38.0	41.8	62.6#
57	10.9	15.0	22.4#
100	6.5	8.9	13.3#

