

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV011520\
 Data File : VV014310.D
 Acq On : 15 Jan 2020 12:45
 Operator : SY/MD
 Sample : VV0115WBL01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK497

Quant Time: Jan 16 03:12:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR011520WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 16 03:02:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev (Min)
1) 1,4-Difluorobenzene	5.66	114	910030	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	827811	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.29	152	376884	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	273799	4.75	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.00%
7) Chloroethane-d5	1.58	69	217028	4.74	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.80%
11) 1,1-Dichloroethene-d2	2.13	63	325513	3.71	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.20%
20) 2-Butanone-d5	3.96	46	666400	48.09	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.18%
24) Chloroform-d	4.40	84	532047	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.40%
26) 1,2-Dichloroethane-d4	5.08	65	273049	4.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.40%
32) Benzene-d6	5.10	84	1175183	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
36) 1,2-Dichloropropane-d6	6.11	67	363170	4.85	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.00%
41) Toluene-d8	7.36	98	1054458	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.60%
43) trans-1,3-Dichloropropene-	7.66	79	141600	4.56	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.20%
46) 2-Hexanone-d5	8.13	63	588609	49.55	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.10%
56) 1,1,2,2-Tetrachloroethane-	10.26	84	236928	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.80%
66) 1,2-Dichlorobenzene-d4	11.67	152	331464	4.91	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.20%

Target Compounds					Ovalue
16) Methylene chloride	2.54	84	27543	0.347 ug/L	98
25) Chloroform	4.42	83	38049	0.341 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV011520\
Data File : VV014310.D
Acq On : 15 Jan 2020 12:45
Operator : SY/MD
Sample : VV0115WBL01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK497

Quant Time: Jan 16 03:12:04 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR011520WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 16 03:02:06 2020
Response via : Initial Calibration

