

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011524\
 Data File : VW033629.D
 Acq On : 15 Jan 2024 09:53
 Operator : SY/MD
 Sample : VW0115WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK394

Quant Time: Jan 16 03:42:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jan 16 03:40:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	131895	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	120152	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	59116	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	31081	4.309	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.200%
7) Chloroethane-d5	1.536	69	29844	4.666	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.400%
11) 1,1-Dichloroethene-d2	2.063	65	14820	4.560	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.200%
20) 2-Butanone-d5	3.831	46	78440	44.482	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.960%
24) Chloroform-d	4.252	84	68388	4.385	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.600%
26) 1,2-Dichloroethane-d4	4.947	65	32731	4.680	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
32) Benzene-d6	4.963	84	136568	4.711	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.200%
36) 1,2-Dichloropropane-d6	5.992	67	44128	4.565	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.400%
41) Toluene-d8	7.246	98	121223	4.684	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
43) trans-1,3-Dichloroprop...	7.561	79	14279	4.273	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.400%
46) 2-Hexanone-d5	8.030	63	54913	41.856	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.720%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	26590	4.708	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.200%
66) 1,2-Dichlorobenzene-d4	11.561	152	43203	4.760	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011524\
Data File : VV033629.D
Acq On : 15 Jan 2024 09:53
Operator : SY/MD
Sample : VV0115WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK394

Quant Time: Jan 16 03:42:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jan 16 03:40:53 2024
Response via : Initial Calibration

