

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV011620\
 Data File : VV014322.D
 Acq On : 16 Jan 2020 16:15
 Operator : SY/MD
 Sample : L1156-05
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jan 17 08:51:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR010220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 17 05:31:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	865517	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	780770	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	365131	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	217527	3.59	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.80%
7) Chloroethane-d5	1.59	69	192591	4.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.20%
11) 1,1-Dichloroethene-d2	2.13	63	277396	3.11	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.20%
20) 2-Butanone-d5	3.95	46	508863	34.98	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	69.96%
24) Chloroform-d	4.40	84	474666	4.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.00%
26) 1,2-Dichloroethane-d4	5.08	65	241613	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.40%
32) Benzene-d6	5.09	84	1018992	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.40%
36) 1,2-Dichloropropane-d6	6.11	67	322926	4.47	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.40%
41) Toluene-d8	7.35	98	909432	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.40%
43) trans-1,3-Dichloropropene-	7.66	79	125490	3.93	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.60%
46) 2-Hexanone-d5	8.13	63	533040	39.40	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	78.80%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	210980	4.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	292790	4.27	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.40%

Target Compounds					Ovalue
16) Methylene chloride	2.54	84	33532	0.533 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV011620\
Data File : VV014322.D
Acq On : 16 Jan 2020 16:15
Operator : SY/MD
Sample : L1156-05
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jan 17 08:51:34 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR010220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 17 05:31:19 2020
Response via : Initial Calibration

