

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV011620\
 Data File : VV014318.D
 Acq On : 16 Jan 2020 13:44
 Operator : SY/MD
 Sample : L1156-01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AA3

Quant Time: Jan 17 05:53:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR010220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 17 05:31:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	914442	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	827712	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	378741	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	204507	3.20	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	64.00%
7) Chloroethane-d5	1.59	69	165874	3.35	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	67.00%
11) 1,1-Dichloroethene-d2	2.13	63	258582	2.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	54.80%#
20) 2-Butanone-d5	3.96	46	595627	38.75	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	77.50%
24) Chloroform-d	4.40	84	451724	3.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	73.00%
26) 1,2-Dichloroethane-d4	5.08	65	225123	3.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.60%
32) Benzene-d6	5.10	84	974981	3.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.00%
36) 1,2-Dichloropropane-d6	6.12	67	306071	4.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.00%
41) Toluene-d8	7.36	98	855370	3.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.80%
43) trans-1,3-Dichloropropene-	7.66	79	122647	3.62	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.40%
46) 2-Hexanone-d5	8.14	63	520083	36.26	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.52%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	206430	3.85	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	77.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	277625	3.90	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	78.00%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV011620\
Data File : VV014318.D
Acq On : 16 Jan 2020 13:44
Operator : SY/MD
Sample : L1156-01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AA3

Quant Time: Jan 17 05:53:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR010220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 17 05:31:19 2020
Response via : Initial Calibration

