

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011623\
 Data File : VW029884.D
 Acq On : 16 Jan 2023 18:23
 Operator : SY/MD
 Sample : 01136-14 10X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGJC7

Quant Time: Jan 17 00:39:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 12 22:05:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	203013	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	194129	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.236	152	106910	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	59087	3.837	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	76.800%	
7) Chloroethane-d5	1.565	69	59237	4.641	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	92.800%	
11) 1,1-Dichloroethene-d2	2.105	63	86998	3.099	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	62.000%	
20) 2-Butanone-d5	3.889	46	166623	61.051	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	122.100%	
24) Chloroform-d	4.339	84	142698	5.052	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.000%	
26) 1,2-Dichloroethane-d4	5.024	65	63032	5.269	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	105.400%	
32) Benzene-d6	5.040	84	303927	5.278	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	105.600%	
36) 1,2-Dichloropropane-d6	6.063	67	94749	5.622	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	112.400%	
41) Toluene-d8	7.307	98	258841	4.855	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.200%	
43) trans-1,3-Dichloroprop...	7.613	79	29053	4.720	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	94.400%	
46) 2-Hexanone-d5	8.085	63	150070	59.889	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	119.780%	
56) 1,1,2,2-Tetrachloroeth...	10.207	84	75573	6.119	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	122.400%#	
66) 1,2-Dichlorobenzene-d4	11.612	152	104542	5.379	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	107.600%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.118	96	2677	0.154	ug/L #	1
14) Carbon disulfide	2.294	76	2922	0.059	ug/L #	93
16) Methylene chloride	2.503	84	11287	0.524	ug/L	88
19) 1,1-Dichloroethane	3.185	63	10607	0.393	ug/L	93
22) cis-1,2-Dichloroethene	3.899	96	2627	0.153	ug/L	93
25) Chloroform	4.368	83	6636	0.226	ug/L	98
34) Trichloroethene	5.908	95	8411	0.517	ug/L	95
47) Tetrachloroethene	7.966	164	950	0.069	ug/L #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011623\
Data File : VV029884.D
Acq On : 16 Jan 2023 18:23
Operator : SY/MD
Sample : 01136-14 10X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGJC7

Quant Time: Jan 17 00:39:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 12 22:05:54 2023
Response via : Initial Calibration

