

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011623\
 Data File : VW029886.D
 Acq On : 16 Jan 2023 19:33
 Operator : SY/MD
 Sample : 01136-21
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Jan 17 00:39:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 12 22:05:54 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	183105	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	177666	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	91815	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	96197	6.927	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 138.600%#		
7) Chloroethane-d5	1.564	69	82028	7.125	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 142.400%#		
11) 1,1-Dichloroethene-d2	2.104	63	131596	5.198	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 104.000%		
20) 2-Butanone-d5	3.902	46	161472	65.596	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 131.200%#		
24) Chloroform-d	4.339	84	163939	6.435	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 128.600%#		
26) 1,2-Dichloroethane-d4	5.024	65	70446	6.529	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 130.600%#		
32) Benzene-d6	5.040	84	355327	6.742	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 134.800%#		
36) 1,2-Dichloropropane-d6	6.063	67	103050	6.681	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 133.600%		
41) Toluene-d8	7.307	98	308519	6.323	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 126.400%		
43) trans-1,3-Dichloroprop...	7.616	79	30913	5.488	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 109.800%		
46) 2-Hexanone-d5	8.085	63	148285	64.660	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 129.320%		
56) 1,1,2,2-Tetrachloroeth...	10.204	84	70694	6.254	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 125.000%#		
66) 1,2-Dichlorobenzene-d4	11.612	152	115773	6.936	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 138.800%#		
Target Compounds						Qvalue
16) Methylene chloride	2.500	84	7214	0.371	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011623\
Data File : VV029886.D
Acq On : 16 Jan 2023 19:33
Operator : SY/MD
Sample : 01136-21
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Jan 17 00:39:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 12 22:05:54 2023
Response via : Initial Calibration

