

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV011720\
 Data File : VV014326.D
 Acq On : 17 Jan 2020 10:11
 Operator : SY/MD
 Sample : VV0117WBL01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK51

Quant Time: Jan 17 23:01:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR010220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 17 22:51:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	815863	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	750321	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	345901	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	231355	4.05	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.00%
7) Chloroethane-d5	1.58	69	187803	4.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.20%
11) 1,1-Dichloroethene-d2	2.13	63	282351	3.35	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.00%
20) 2-Butanone-d5	3.95	46	564090	41.14	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.28%
24) Chloroform-d	4.40	84	503076	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.20%
26) 1,2-Dichloroethane-d4	5.08	65	255264	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.40%
32) Benzene-d6	5.10	84	1093260	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.60%
36) 1,2-Dichloropropane-d6	6.11	67	344680	4.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.40%
41) Toluene-d8	7.36	98	969944	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.80%
43) trans-1,3-Dichloropropene-	7.66	79	134830	4.39	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.80%
46) 2-Hexanone-d5	8.13	63	587052	45.15	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.30%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	232707	4.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	315830	4.86	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.20%

Target Compounds					Ovalue
16) Methylene chloride	2.53	84	13445	0.227 ug/L	95
25) Chloroform	4.42	83	19212	0.187 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV011720\
Data File : VV014326.D
Acq On : 17 Jan 2020 10:11
Operator : SY/MD
Sample : VV0117WBL01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK51

Quant Time: Jan 17 23:01:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR010220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 17 22:51:36 2020
Response via : Initial Calibration

