

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011724\
 Data File : VV033707.D
 Acq On : 17 Jan 2024 10:04
 Operator : SY/MD
 Sample : VV0117WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK397

Quant Time: Jan 18 00:03:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 18 00:02:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	109412	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	101066	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	48088	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	19851	3.318	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.400%
7) Chloroethane-d5	1.535	69	21378	4.029	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.600%
11) 1,1-Dichloroethene-d2	2.063	65	9740	3.612	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.200%
20) 2-Butanone-d5	3.834	46	61353	41.941	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.880%
24) Chloroform-d	4.252	84	55069	4.256	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.200%
26) 1,2-Dichloroethane-d4	4.947	65	26864	4.631	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.600%
32) Benzene-d6	4.963	84	103124	4.229	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.600%
36) 1,2-Dichloropropane-d6	5.992	67	33422	4.111	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	82.200%
41) Toluene-d8	7.246	98	89652	4.118	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.400%
43) trans-1,3-Dichloroprop...	7.561	79	11151	3.967	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.400%
46) 2-Hexanone-d5	8.030	63	47679	43.205	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.420%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	22123	4.657	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.200%
66) 1,2-Dichlorobenzene-d4	11.561	152	33669	4.560	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.200%
Target Compounds						
16) Methylene chloride	2.455	84	2443	0.268	ug/L	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011724\
Data File : VV033707.D
Acq On : 17 Jan 2024 10:04
Operator : SY/MD
Sample : VV0117WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK397

Quant Time: Jan 18 00:03:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 18 00:02:20 2024
Response via : Initial Calibration

