




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration



TIC: VV038446.D\data.ms

(13) Acetone (T)

2.143min (+ 0.003) 2.02 ug/L

response 5285

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	33.60	81.38#
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(13) Acetone (T)

2.143min (+ 0.003) 5.74 ug/L m

response 14997

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	33.60	28.68
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



(15) Methyl Acetate (T)

2.410min (+ 0.026) 0.47 ug/L

response 3167

Ion	Exp%	Act%
43.05	100.00	100.00
74.05	25.50	11.71#
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(15) Methyl Acetate (T)

2.410min (+ 0.026) 0.57 ug/L m

response 3809

Ion	Exp%	Act%
43.05	100.00	100.00
74.05	25.50	9.74#
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration



TIC: VV038446.D\data.ms

(20) 2-Butanone-d5 (S)**3.899min (+ 0.097) 0.93 ug/L****response 2962**

Ion	Exp%	Act%
46.10	100.00	100.00
77.10	24.70	26.47
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration



TIC: VV038446.D\data.ms

(20) 2-Butanone-d5 (S)**3.921min (+ 0.119) 4.15 ug/L m****response 13264**

Ion	Exp%	Act%
46.10	100.00	100.00
77.10	24.70	5.91#
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(21) 2-Butanone (T)

4.002min (+ 0.122) 0.50 ug/L

response 1733

Ion	Exp%	Act%
43.05	100.00	100.00
72.10	24.30	15.23
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(21) 2-Butanone (T)

3.966min (+ 0.087) 4.71 ug/L m

response 16178

Ion	Exp%	Act%
43.05	100.00	100.00
72.10	24.30	1.63#
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(26) 1,2-Dichloroethane-d4 (S)

4.960min (+ 0.026) 0.30 ug/L

response 4067

Ion	Exp%	Act%
65.10	100.00	100.00
67.05	53.30	41.55
51.05	35.10	40.05
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(26) 1,2-Dichloroethane-d4 (S)

4.960min (+ 0.026) 0.48 ug/L m

response 6627

Ion	Exp%	Act%
65.10	100.00	100.00
67.05	53.30	25.50#
51.05	35.10	24.58
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration



TIC: VV038446.D\data.ms

(27) 1,2-Dichloroethane (T)

5.059min (+ 0.026) 0.45 ug/L

response 7871

Ion	Exp%	Act%
62.00	100.00	100.00
98.05	9.70	11.95#
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration



TIC: VV038446.D\data.ms

(27) 1,2-Dichloroethane (T)

5.059min (+ 0.026) 0.50 ug/L m

response 8660

Ion	Exp%	Act%
62.00	100.00	100.00
98.05	9.70	11.95#
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(36) 1,2-Dichloropropane-d6 (S)

5.995min (+ 0.016) 0.51 ug/L

response 9025

Ion	Exp%	Act%
67.05	100.00	100.00
65.10	68.50	59.15
46.10	48.90	45.07
42.05	26.00	18.58




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(36) 1,2-Dichloropropane-d6 (S)

5.995min (+ 0.016) 0.42 ug/L m

response 7473

Ion	Exp%	Act%
67.05	100.00	100.00
65.10	68.50	71.43
46.10	48.90	54.44
42.05	26.00	22.44




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration



TIC: VV038446.D\data.ms

(37) 1,2-Dichloropropane (T)

6.101min (+ 0.019) 0.42 ug/L

response 7063

Ion	Exp%	Act%
63.10	100.00	100.00
112.10	5.10	2.04#
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(37) 1,2-Dichloropropane (T)

6.101min (+ 0.019) 0.47 ug/L m

response 7819

Ion	Exp%	Act%
63.10	100.00	100.00
112.10	5.10	1.84#
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(40) 4-Methyl-2-pentanone (T)

7.169min (+ 0.019) 3.50 ug/L

response 28799

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	39.60	43.61
100.15	15.50	14.91
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration



TIC: VV038446.D\data.ms

(40) 4-Methyl-2-pentanone (T)

7.169min (+ 0.019) 4.32 ug/L m

response 35545

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	39.60	35.33
100.15	15.50	12.08#
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration



TIC: VV038446.D\data.ms

(43) trans-1,3-Dichloropropene-d4 (S)

7.616min (+ 0.068) 0.07 ug/L

response 433

Ion	Exp%	Act%
79.10	100.00	100.00
81.05	29.90	25.64
42.05	34.80	30.02
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
 Data File : VV038446.D
 Acq On : 17 Jan 2025 10:40
 Operator : SY/MD
 Sample : VSTD0.523
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jan 18 01:29:26 2025
 Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
 Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(43) trans-1,3-Dichloropropene-d4 (S)

7.577min (+ 0.029) 0.39 ug/L m

response 2390

Ion	Exp%	Act%
79.10	100.00	100.00
81.05	29.90	4.64#
42.05	34.80	5.44#
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration



TIC: VV038446.D\data.ms

(44) trans-1,3-Dichloropropene (T)

7.677min (+ 0.100) 0.03 ug/L

response 481

Ion	Exp%	Act%
75.05	100.00	100.00
77.05	32.60	28.73
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(44) trans-1,3-Dichloropropene (T)

7.600min (+ 0.023) 0.44 ug/L m

response 8427

Ion	Exp%	Act%
75.05	100.00	100.00
77.05	32.60	22.64#
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration



TIC: VV038446.D\data.ms

(46) 2-Hexanone-d5 (S)

8.063min (+ 0.042) 2.67 ug/L

response 6287

Ion	Exp%	Act%
63.10	100.00	100.00
46.10	172.70	166.95
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(46) 2-Hexanone-d5 (S)

8.063min (+ 0.042) 3.44 ug/L m

response 8119

Ion	Exp%	Act%
63.10	100.00	100.00
46.10	172.70	129.28
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(48) 2-Hexanone (T)

8.098min (+ 0.029) 3.21 ug/L

response 18984

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	54.50	50.73
57.20	18.70	15.42
100.15	12.50	4.98#




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(48) 2-Hexanone (T)

8.098min (+ 0.029) 4.12 ug/L m

response 24378

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	54.50	39.50#
57.20	18.70	12.01#
100.15	12.50	3.88#




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(53) m,p-Xylene (T)

9.075min (+ 0.013) 0.38 ug/L

response 10223

Ion	Exp%	Act%
106.15	100.00	100.00
91.15	211.70	197.13
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(53) m,p-Xylene (T)

9.075min (+ 0.013) 0.44 ug/L m

response 11867

Ion	Exp%	Act%
106.15	100.00	100.00
91.15	211.70	197.13
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(55) Styrene (T)

9.500min (+ 0.013) 0.37 ug/L

response 16580

Ion	Exp%	Act%
104.10	100.00	100.00
78.10	44.00	44.23
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(55) Styrene (T)

9.500min (+ 0.013) 0.39 ug/L m

response 17869

Ion	Exp%	Act%
104.10	100.00	100.00
78.10	44.00	44.23
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration



TIC: VV038446.D\data.ms

(70) 1,2,4-trichlorobenzene (T)

13.201min (+ 0.010) 0.41 ug/L

response 6350

Ion	Exp%	Act%
179.90	100.00	100.00
182.00	96.90	101.98
144.95	32.80	38.96
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025



TIC: VV038446.D\data.ms

(70) 1,2,4-trichlorobenzene (T)

13.201min (+ 0.010) 0.43 ug/L m

response 6657

Ion	Exp%	Act%
179.90	100.00	100.00
182.00	96.90	97.28
144.95	32.80	37.16
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration



TIC: VV038446.D\data.ms

(71) Naphthalene

13.451min (+ 0.019) 0.32 ug/L

response 7341

Ion	Exp%	Act%
128.00	100.00	100.00
129.00	10.40	10.64
127.00	13.80	12.06
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
Data File : VV038446.D
Acq On : 17 Jan 2025 10:40
Operator : SY/MD
Sample : VSTD0.523
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5223

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 01/20/2025
Supervised By :Semsettin Yesilyurt 01/21/2025

Quant Time: Jan 18 01:40:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 18 01:29:26 2025
Response via : Initial Calibration



TIC: VV038446.D\data.ms

(71) Naphthalene

13.451min (+ 0.019) 0.44 ug/L m

response 9945

Ion	Exp%	Act%
128.00	100.00	100.00
129.00	10.40	7.85#
127.00	13.80	8.90#
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VW011725\
 Data File : VW038446.D
 Acq On : 17 Jan 2025 10:40
 Operator : SY/MD
 Sample : VSTD0.523
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5223

Manual IntegrationsAPPROVED

Quant Time: Jan 18 01:40:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jan 18 01:29:26 2025
 Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/20/2025
 Supervised By :Semsettin Yesilyurt 01/21/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	194122	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	204099	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	92206	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	7140	0.446	ug/L	0.00
7) Chloroethane-d5	1.532	69	6314	0.470	ug/L	0.00
11) 1,1-Dichloroethene-d2	2.056	65	3559	0.493	ug/L	0.00
20) 2-Butanone-d5	3.921	46	13264m	4.149	ug/L	0.12
24) Chloroform-d	4.252	84	13927	0.464	ug/L	0.02
26) 1,2-Dichloroethane-d4	4.960	65	6627m	0.482	ug/L	0.03
32) Benzene-d6	4.969	84	22648	0.403	ug/L	0.02
36) 1,2-Dichloropropane-d6	5.995	67	7473m	0.422	ug/L	0.02
41) Toluene-d8	7.246	98	19032	0.377	ug/L	0.01
43) trans-1,3-Dichloroprop...	7.577	79	2390m	0.386	ug/L	0.03
46) 2-Hexanone-d5	8.063	63	8119m	3.445	ug/L	0.04
56) 1,1,2,2-Tetrachloroeth...	10.149	84	5656	0.418	ug/L	0.00
66) 1,2-Dichlorobenzene-d4	11.561	152	7072	0.453	ug/L	0.00
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.105	85	11771	0.541	ug/L	97
3) Chloromethane	1.214	50	12582	0.566	ug/L	95
5) Vinyl chloride	1.281	62	11890	0.529	ug/L	100
6) Bromomethane	1.487	94	6159	0.548	ug/L	92
8) Chloroethane	1.548	64	7044	0.482	ug/L	96
9) Trichlorofluoromethane	1.712	101	15730	0.550	ug/L	99
10) 1,1,2-Trichloro-1,2,2-...	2.066	101	8384	0.513	ug/L #	91
12) 1,1-Dichloroethene	2.066	96	8816	0.548	ug/L	91
13) Acetone	2.143	43	14997m	5.735	ug/L	
14) Carbon disulfide	2.236	76	28507	0.527	ug/L	99
15) Methyl Acetate	2.410	43	3809m	0.570	ug/L	
16) Methylene chloride	2.445	84	9876	0.554	ug/L	95
17) Methyl tert-butyl Ether	2.699	73	16122	0.476	ug/L	96
18) trans-1,2-Dichloroethene	2.699	96	8538	0.516	ug/L	93
19) 1,1-Dichloroethane	3.108	63	16439	0.522	ug/L	90
21) 2-Butanone	3.966	43	16178m	4.713	ug/L	
22) cis-1,2-Dichloroethene	3.828	96	8501	0.505	ug/L	86
23) Bromochloromethane	4.156	128	3756	0.491	ug/L	88
25) Chloroform	4.275	83	16215	0.521	ug/L	94
27) 1,2-Dichloroethane	5.059	62	8660m	0.500	ug/L	
29) 1,1,1-Trichloroethane	4.503	97	14150	0.521	ug/L	98
30) Cyclohexane	4.577	56	10922	0.442	ug/L	95
31) Carbon tetrachloride	4.728	117	12453	0.519	ug/L	98
33) Benzene	5.014	78	29809	0.460	ug/L	100
34) Trichloroethene	5.841	95	8218	0.488	ug/L	93
35) Methylcyclohexane	6.043	83	11564	0.450	ug/L	94
37) 1,2-Dichloropropane	6.101	63	7819m	0.467	ug/L	
38) Bromodichloromethane	6.439	83	10654	0.491	ug/L #	94
39) cis-1,3-Dichloropropene	6.966	75	9741	0.422	ug/L	100
40) 4-Methyl-2-pentanone	7.169	43	35545m	4.324	ug/L	
42) Toluene	7.317	91	27733	0.411	ug/L	99
44) trans-1,3-Dichloropropene	7.600	75	8427m	0.438	ug/L	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV011725\
 Data File : VV038446.D
 Acq On : 17 Jan 2025 10:40
 Operator : SY/MD
 Sample : VSTD0.523
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5223

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 01/20/2025
 Supervised By :Semsettin Yesilyurt 01/21/2025

Quant Time: Jan 18 01:40:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jan 18 01:29:26 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

45) 1,1,2-Trichloroethane	7.770	97	5766	0.479	ug/L	85
47) Tetrachloroethene	7.905	164	6402	0.491	ug/L	96
48) 2-Hexanone	8.098	43	24378m	4.120	ug/L	
49) Dibromochloromethane	8.175	129	6382	0.461	ug/L	74
50) 1,2-Dibromoethane	8.288	107	5080	0.454	ug/L #	90
51) Chlorobenzene	8.812	112	22448	0.494	ug/L	98
52) Ethylbenzene	8.947	91	31424	0.432	ug/L	99
53) m,p-Xylene	9.075	106	11867m	0.441	ug/L	
54) o-Xylene	9.477	106	10450	0.398	ug/L	98
55) Styrene	9.500	104	17869m	0.395	ug/L	
57) 1,1,2,2-Tetrachloroethane	10.175	83	7233	0.493	ug/L	90
59) Bromoform	9.670	173	3287	0.511	ug/L #	98
60) Isopropylbenzene	9.860	105	28144	0.454	ug/L	99
61) 1,2,3-Trichloropropane	10.207	75	4777	0.539	ug/L	96
62) 1,3,5-Trimethylbenzene	10.471	105	20919	0.431	ug/L	100
63) 1,2,4-Trimethylbenzene	10.847	105	20385	0.420	ug/L	96
64) 1,3-Dichlorobenzene	11.120	146	15260	0.496	ug/L	99
65) 1,4-Dichlorobenzene	11.204	146	17508	0.550	ug/L	97
67) 1,2-Dichlorobenzene	11.577	146	14453	0.501	ug/L	95
68) 1,2-Dibromo-3-chloropr...	12.365	75	895	0.489	ug/L	85
69) 1,3,5-Trichlorobenzene	12.580	180	10770	0.509	ug/L	96
70) 1,2,4-trichlorobenzene	13.201	180	6657m	0.426	ug/L	
71) Naphthalene	13.451	128	9945m	0.440	ug/L	
72) 1,2,3-Trichlorobenzene	13.680	180	5669	0.417	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

