

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011821\
 Data File : VW020042.D
 Acq On : 18 Jan 2021 14:21
 Operator : SY/MD
 Sample : M1147-15 100X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG4Y7

Quant Time: Jan 19 02:49:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM011121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jan 19 02:46:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.621	114	379598	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.858	117	369186	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.257	152	218392	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.306	65	125398	39.16	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	78.32%
7) Chloroethane-d5	1.569	69	116655	45.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.28%
11) 1,1-Dichloroethene-d2	2.110	63	186011	29.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.00%#
21) 2-Butanone-d5	3.897	46	211697	101.92	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.92%
24) Chloroform-d	4.354	84	255450	44.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.58%
26) 1,2-Dichloroethane-d4	5.039	65	186350	47.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.06%
32) Benzene-d6	5.055	84	461191	44.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.52%
36) 1,2-Dichloropropane-d6	6.074	67	146739	45.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.22%
41) Toluene-d8	7.322	98	406286	43.02	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.04%
43) trans-1,3-Dichloroprop...	7.627	79	69605	41.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.46%
47) 2-Hexanone-d5	8.093	63	115645	88.67	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.67%
57) 1,1,2,2-Tetrachloroeth...	10.222	84	227269	51.68	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	103.36%
64) 1,2-Dichlorobenzene-d4	11.633	152	202054	47.90	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.80%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011821\
Data File : VV020042.D
Acq On : 18 Jan 2021 14:21
Operator : SY/MD
Sample : M1147-15 100X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG4Y7

Quant Time: Jan 19 02:49:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM011121WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jan 19 02:46:41 2021
Response via : Initial Calibration

