

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011824\
 Data File : VW033730.D
 Acq On : 18 Jan 2024 10:40
 Operator : SY/MD
 Sample : VW0118WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK398

Quant Time: Jan 19 00:09:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 19 00:08:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	113890	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	104478	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	50445	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	35409	5.685	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	113.800%
7) Chloroethane-d5	1.535	69	28612	5.180	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.600%
11) 1,1-Dichloroethene-d2	2.063	65	16408	5.846	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	117.000%
20) 2-Butanone-d5	3.847	46	47388	31.121	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	62.240%
24) Chloroform-d	4.252	84	60823	4.516	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.400%
26) 1,2-Dichloroethane-d4	4.950	65	26706	4.423	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.400%
32) Benzene-d6	4.963	84	126257	5.008	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.200%
36) 1,2-Dichloropropane-d6	5.992	67	34993	4.164	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.200%
41) Toluene-d8	7.246	98	117591	5.225	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.400%
43) trans-1,3-Dichloroprop...	7.561	79	12324	4.241	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.800%
46) 2-Hexanone-d5	8.034	63	36530	32.021	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	64.040%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	18687	3.805	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	76.200%
66) 1,2-Dichlorobenzene-d4	11.561	152	34696	4.480	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011824\
Data File : VV033730.D
Acq On : 18 Jan 2024 10:40
Operator : SY/MD
Sample : VV0118WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK398

Quant Time: Jan 19 00:09:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 19 00:08:23 2024
Response via : Initial Calibration

