

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011824\
 Data File : VV033737.D
 Acq On : 18 Jan 2024 13:57
 Operator : SY/MD
 Sample : P1143-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Jan 19 00:11:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 19 00:08:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	109030	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	101145	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	49991	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	35387	5.935	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	118.800%
7) Chloroethane-d5	1.536	69	28761	5.439	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.800%
11) 1,1-Dichloroethene-d2	2.060	65	15992	5.952	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	119.000%
20) 2-Butanone-d5	3.815	46	41692	28.601	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	57.200%
24) Chloroform-d	4.249	84	59048	4.580	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
26) 1,2-Dichloroethane-d4	4.944	65	26439	4.573	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.400%
32) Benzene-d6	4.960	84	124403	5.097	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.000%
36) 1,2-Dichloropropane-d6	5.989	67	33766	4.150	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.000%
41) Toluene-d8	7.243	98	115603	5.306	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.200%
43) trans-1,3-Dichloroprop...	7.558	79	12083	4.295	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.000%
46) 2-Hexanone-d5	8.031	63	31894	28.879	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	57.760%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	17211	3.620	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	72.400%
66) 1,2-Dichlorobenzene-d4	11.561	152	34653	4.515	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.200%
Target Compounds						Qvalue
16) Methylene chloride	2.449	84	8396	0.926	ug/L	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011824\
Data File : VV033737.D
Acq On : 18 Jan 2024 13:57
Operator : SY/MD
Sample : P1143-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Jan 19 00:11:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 19 00:08:23 2024
Response via : Initial Calibration





#16
Methylene chloride
Concen: 0.926 ug/L
RT: 2.449 min Scan# 41
Delta R.T. 0.000 min
Lab File: VV033737.D
Acq: 18 Jan 2024 13:57

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Tgt Ion:	84	Resp:	8396
Ion	Ratio	Lower	Upper
84	100		
86	64.4	45.5	84.5
49	143.0	100.3	186.3

