

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011924\
 Data File : VW033748.D
 Acq On : 19 Jan 2024 12:51
 Operator : SY/MD
 Sample : P1190-04 100X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCMP7

Quant Time: Jan 19 22:22:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 19 22:20:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	253671	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	241552	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	121246	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	78864	40.452	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.900%
7) Chloroethane-d5	1.532	69	66978	43.382	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.760%
11) 1,1-Dichloroethene-d2	2.060	65	37563	44.769	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.540%
21) 2-Butanone-d5	3.789	46	100985	75.518	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	75.520%
24) Chloroform-d	4.249	84	147755	42.384	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.760%
26) 1,2-Dichloroethane-d4	4.941	65	100919	44.804	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.600%
32) Benzene-d6	4.957	84	296130	43.123	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.240%
36) 1,2-Dichloropropane-d6	5.986	67	94351	41.740	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.480%
41) Toluene-d8	7.243	98	271056	43.808	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.620%
43) trans-1,3-Dichloroprop...	7.551	79	46307	41.451	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.900%
47) 2-Hexanone-d5	8.024	63	72694	77.530	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	77.530%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	118370	41.019	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.040%
66) 1,2-Dichlorobenzene-d4	11.561	152	107373	46.676	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011924\
Data File : VV033748.D
Acq On : 19 Jan 2024 12:51
Operator : SY/MD
Sample : P1190-04 100X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCMP7

Quant Time: Jan 19 22:22:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 19 22:20:26 2024
Response via : Initial Calibration

