

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011924\
 Data File : VW033752.D
 Acq On : 19 Jan 2024 14:26
 Operator : SY/MD
 Sample : P1190-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCMP6

Quant Time: Jan 19 22:23:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 19 22:20:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	245734	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	233240	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	115887	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	71623	37.924	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.840%
7) Chloroethane-d5	1.532	69	61291	40.980	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.960%
11) 1,1-Dichloroethene-d2	2.059	65	34274	42.168	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.340%
21) 2-Butanone-d5	3.783	46	89562	69.139	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	69.140%
24) Chloroform-d	4.249	84	139910	41.430	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.860%
26) 1,2-Dichloroethane-d4	4.940	65	97002	44.456	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.920%
32) Benzene-d6	4.960	84	281354	42.431	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.860%
36) 1,2-Dichloropropane-d6	5.989	67	89483	40.997	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	82.000%
41) Toluene-d8	7.243	98	257524	43.104	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.200%
43) trans-1,3-Dichloroprop...	7.554	79	43978	40.769	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.540%
47) 2-Hexanone-d5	8.024	63	64139	70.843	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	70.840%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	112300	40.302	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	80.600%
66) 1,2-Dichlorobenzene-d4	11.561	152	97502	44.345	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011924\
Data File : VV033752.D
Acq On : 19 Jan 2024 14:26
Operator : SY/MD
Sample : P1190-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCMP6

Quant Time: Jan 19 22:23:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 19 22:20:26 2024
Response via : Initial Calibration

