

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011924\
 Data File : VW033760.D
 Acq On : 19 Jan 2024 17:37
 Operator : SY/MD
 Sample : P1190-11
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCMQ9

Quant Time: Jan 19 22:25:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 19 22:20:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	190400	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	179856	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	86260	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	51340	35.085	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	70.160%
7) Chloroethane-d5	1.529	69	46132	39.809	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.620%
11) 1,1-Dichloroethene-d2	2.060	65	24648	39.138	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.280%
21) 2-Butanone-d5	3.783	46	83270	82.963	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	82.960%
24) Chloroform-d	4.249	84	110642	42.285	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.560%
26) 1,2-Dichloroethane-d4	4.941	65	79827	47.217	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.440%
32) Benzene-d6	4.960	84	215571	42.160	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.320%
36) 1,2-Dichloropropane-d6	5.989	67	71109	42.249	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	84.500%
41) Toluene-d8	7.243	98	194619	42.244	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.480%
43) trans-1,3-Dichloroprop...	7.555	79	32702	39.314	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.620%
47) 2-Hexanone-d5	8.024	63	60183	86.204	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.200%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	95165	44.290	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.580%
66) 1,2-Dichlorobenzene-d4	11.561	152	75930	46.395	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011924\
Data File : VV033760.D
Acq On : 19 Jan 2024 17:37
Operator : SY/MD
Sample : P1190-11
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCMQ9

Quant Time: Jan 19 22:25:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 19 22:20:26 2024
Response via : Initial Calibration

