

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011924\
 Data File : VW033774.D
 Acq On : 19 Jan 2024 23:58
 Operator : SY/MD
 Sample : P1189-12
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCMP5

Quant Time: Jan 20 00:32:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 19 23:43:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	235522	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	219294	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	109382	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	63843	35.270	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	70.540%
7) Chloroethane-d5	1.532	69	53794	37.527	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	75.060%
11) 1,1-Dichloroethene-d2	2.059	65	30367	38.982	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.960%
21) 2-Butanone-d5	3.783	46	78761	63.437	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	63.440%
24) Chloroform-d	4.249	84	127756	39.471	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.940%
26) 1,2-Dichloroethane-d4	4.940	65	87931	42.046	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.100%
32) Benzene-d6	4.960	84	251173	40.288	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.580%
36) 1,2-Dichloropropane-d6	5.989	67	80869	39.407	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	78.820%
41) Toluene-d8	7.243	98	227740	40.543	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.080%
43) trans-1,3-Dichloroprop...	7.554	79	35340	34.845	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.680%
47) 2-Hexanone-d5	8.024	63	61565	72.325	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	72.320%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	102656	39.184	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	78.360%
66) 1,2-Dichlorobenzene-d4	11.561	152	89236	42.999	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011924\
Data File : VV033774.D
Acq On : 19 Jan 2024 23:58
Operator : SY/MD
Sample : P1189-12
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCMP5

Quant Time: Jan 20 00:32:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 19 23:43:59 2024
Response via : Initial Calibration

