

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012021\
 Data File : VV020079.D
 Acq On : 20 Jan 2021 16:32
 Operator : SY/MD
 Sample : VV0120WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK84

Quant Time: Jan 20 23:40:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM012021WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 20 23:39:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.621	114	362808	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.859	117	351886	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.257	152	190387	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.309	65	103284	45.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.76%
7) Chloroethane-d5	1.569	69	80383	45.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.62%
11) 1,1-Dichloroethene-d2	2.110	63	160082	35.21	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.42%
21) 2-Butanone-d5	3.891	46	152986	96.45	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.45%
24) Chloroform-d	4.354	84	230659	45.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.76%
26) 1,2-Dichloroethane-d4	5.039	65	160369	47.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.24%
32) Benzene-d6	5.055	84	440307	47.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.64%
36) 1,2-Dichloropropane-d6	6.074	67	130153	47.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.74%
41) Toluene-d8	7.322	98	411479	46.99	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.98%
43) trans-1,3-Dichloroprop...	7.627	79	68399	44.94	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.88%
47) 2-Hexanone-d5	8.093	63	96287	85.76	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.76%
57) 1,1,2,2-Tetrachloroeth...	10.222	84	169858	44.42	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.84%
64) 1,2-Dichlorobenzene-d4	11.633	152	182421	48.95	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.90%
Target Compounds						
48) 2-Hexanone	8.145	43	8657	3.40	ug/L	Qvalue # 87


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012021\
Data File : VV020079.D
Acq On : 20 Jan 2021 16:32
Operator : SY/MD
Sample : VV0120WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK84

Quant Time: Jan 20 23:40:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM012021WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 20 23:39:20 2021
Response via : Initial Calibration





#48
2-Hexanone
Concen: 3.40 ug/L
RT: 8.145 min Scan# 2194
Delta R.T. 0.000 min
Lab File: VV020079.D
Acq: 20 Jan 2021 16:32

Instrument :
MSVOA_V
ClientSampleId :
VBLK84

Tgt Ion: 43	Resp: 8657
Ion Ratio	Lower Upper
43	100
58	42.5 41.8 62.6
57	13.4 15.0 22.4#
100	7.0 8.9 13.3#

